


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF031519\
 Data File : BF113059.D
 Acq On : 15 Mar 2019 11:38
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF022719.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Mar 13 03:42:55 2019



AutoFind: Scans 1600, 1601, 1602; Background Corrected with Scan 1595

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.4	59872	PASS
68	69	0.00	2	1.7	928	PASS
69	198	0.00	100	34.8	55824	PASS
70	69	0.00	2	0.5	304	PASS
127	198	10	80	47.1	75458	PASS
197	198	0.00	2	0.5	753	PASS
198	198	100	100	100.0	160269	PASS
199	198	5	9	6.7	10706	PASS
275	198	10	60	27.2	43613	PASS
365	198	1	100	3.2	5133	PASS
441	198	0.01	100	12.3	19766	PASS
442	198	50	100	77.7	124544	PASS
443	442	15	24	19.4	24111	PASS