

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/17/2023	BNA_F	<u>BF132823.D</u>
Compound Name	Response	Retention Time
DDT	682357	13.657
DDD	16184	13.357
DDE	569	13.015
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
16753	699110	2.40