

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF031819\  
 Data File : BF113098.D  
 Acq On : 18 Mar 2019 18:49  
 Operator : JU/SJ  
 Sample : K1687-07 5X  
 Misc :  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 HR-05-031519-A

Quant Time: Mar 19 05:38:53 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Mar 13 03:42:55 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.73  | 152  | 153789   | 20.00 | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.01  | 136  | 637737   | 20.00 | ng    | 0.00     |
| 39) Acenaphthene-d10        | 9.76  | 164  | 306015   | 20.00 | ng    | -0.01    |
| 64) Phenanthrene-d10        | 11.24 | 188  | 555972   | 20.00 | ng    | 0.00     |
| 76) Chrysene-d12            | 13.87 | 240  | 488383   | 20.00 | ng    | 0.00     |
| 87) Perylene-d12            | 15.28 | 264  | 410241   | 20.00 | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |       |       |          |
| 5) 2-Fluorophenol           | 5.34  | 112  | 216702   | 21.92 | ng    | 0.00     |
| 7) Phenol-d6                | 6.37  | 99   | 302646   | 24.37 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.29  | 82   | 177213   | 16.63 | ng    | -0.01    |
| 42) 2,4,6-Tribromophenol    | 10.55 | 330  | 84411    | 25.98 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 9.08  | 172  | 419056   | 17.79 | ng    | -0.01    |
| 79) Terphenyl-d14           | 12.82 | 244  | 390208   | 15.49 | ng    | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF031819\  
Data File : BF113098.D  
Acq On : 18 Mar 2019 18:49  
Operator : JU/SJ  
Sample : K1687-07 5X  
Misc :  
ALS Vial : 16 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
HR-05-031519-A

Quant Time: Mar 19 05:38:53 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Mar 13 03:42:55 2019  
Response via : Initial Calibration

