

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031822\
 Data File : BF127390.D
 Acq On : 18 Mar 2022 14:20
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 19 01:51:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 19 01:50:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	110	0.00
2	1,4-Dioxane	40.000	42.870	-7.2	124	0.00
3	Pyridine	40.000	45.072	-12.7	122	0.00
4	n-Nitrosodimethylamine	40.000	41.842	-4.6	112	0.00
5 S	2-Fluorophenol	80.000	82.567	-3.2	118	0.00
6	Aniline	40.000	38.096	4.8	102	0.00
7 S	Phenol-d6	80.000	75.612	5.5	105	0.00
8	2-Chlorophenol	40.000	39.167	2.1	107	0.00
9	Benzaldehyde	40.000	36.737	8.2	102	0.00
10 C	Phenol	40.000	36.862	7.8	100	0.00
11	bis(2-Chloroethyl)ether	40.000	39.010	2.5	104	0.00
12	1,3-Dichlorobenzene	40.000	38.787	3.0	107	0.00
13 C	1,4-Dichlorobenzene	40.000	37.563	6.1	102	0.00
14	1,2-Dichlorobenzene	40.000	37.440	6.4	103	0.00
15	Benzyl Alcohol	40.000	39.683	0.8	105	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.990	10.0	96	0.00
17	2-Methylphenol	40.000	37.903	5.2	104	0.00
18	Hexachloroethane	40.000	36.403	9.0	101	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.129	7.2	99	0.00
20	3+4-Methylphenols	40.000	34.802	13.0	97	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	107	0.00
22	Acetophenone	40.000	37.466	6.3	100	0.00
23 S	Nitrobenzene-d5	80.000	76.647	4.2	102	0.00
24	Nitrobenzene	40.000	40.267	-0.7	106	0.00
25	Isophorone	40.000	41.623	-4.1	112	0.00
26 C	2-Nitrophenol	40.000	41.614	-4.0	112	0.00
27	2,4-Dimethylphenol	40.000	41.345	-3.4	110	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.220	-0.5	104	0.00
29 C	2,4-Dichlorophenol	40.000	39.803	0.5	103	0.00
30	1,2,4-Trichlorobenzene	40.000	39.930	0.2	106	0.00
31	Naphthalene	40.000	39.430	1.4	107	0.00
32	Benzoic acid	40.000	35.072	12.3	96	0.00
33	4-Chloroaniline	40.000	39.945	0.1	108	0.00
34 C	Hexachlorobutadiene	40.000	42.513	-6.3	111	0.00
35	Caprolactam	40.000	39.590	1.0	98	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.055	2.4	101	0.00
37	2-Methylnaphthalene	40.000	39.091	2.3	104	0.00
38	1-Methylnaphthalene	40.000	39.162	2.1	102	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.710	0.7	106	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.700	0.7	102	0.00
42 S	2,4,6-Tribromophenol	80.000	80.266	-0.3	105	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.687	0.8	102	0.00
44	2,4,5-Trichlorophenol	40.000	39.309	1.7	103	0.00
45 S	2-Fluorobiphenyl	80.000	76.596	4.3	105	0.00
46	1,1'-Biphenyl	40.000	37.826	5.4	103	0.00
47	2-Chloronaphthalene	40.000	37.590	6.0	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031822\
 Data File : BF127390.D
 Acq On : 18 Mar 2022 14:20
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 19 01:51:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 19 01:50:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.862	0.3	104	0.00
49	Acenaphthylene	40.000	37.759	5.6	99	0.00
50	Dimethylphthalate	40.000	40.034	-0.1	109	0.00
51	2,6-Dinitrotoluene	40.000	40.175	-0.4	105	0.00
52 C	Acenaphthene	40.000	37.545	6.1	101	0.00
53	3-Nitroaniline	40.000	36.382	9.0	97	0.00
54 P	2,4-Dinitrophenol	40.000	40.457	-1.1	103	0.00
55	Dibenzofuran	40.000	37.374	6.6	100	0.00
56 P	4-Nitrophenol	40.000	50.512	-26.3#	122	0.00
57	2,4-Dinitrotoluene	40.000	38.097	4.8	98	0.00
58	Fluorene	40.000	37.399	6.5	99	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.030	-5.1	109	0.00
60	Diethylphthalate	40.000	38.947	2.6	105	0.00
61	4-Chlorophenyl-phenylether	40.000	37.912	5.2	99	0.00
62	4-Nitroaniline	40.000	39.271	1.8	102	0.00
63	Azobenzene	40.000	38.571	3.6	99	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	108	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.087	-5.2	102	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.560	6.1	102	0.00
67	4-Bromophenyl-phenylether	40.000	37.784	5.5	99	0.00
68	Hexachlorobenzene	40.000	38.586	3.5	100	0.00
69	Atrazine	40.000	38.384	4.0	98	0.00
70 C	Pentachlorophenol	40.000	41.530	-3.8	102	0.00
71	Phenanthrene	40.000	37.206	7.0	103	0.00
72	Anthracene	40.000	39.254	1.9	108	0.00
73	Carbazole	40.000	36.911	7.7	100	0.00
74	Di-n-butylphthalate	40.000	38.524	3.7	105	0.00
75 C	Fluoranthene	40.000	36.372	9.1	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	93	0.00
77	Benzydine	40.000	39.056	2.4	107	0.00
78	Pyrene	40.000	42.864	-7.2	101	0.00
79 S	Terphenyl-d14	80.000	85.612	-7.0	102	0.00
80	Butylbenzylphthalate	40.000	44.496	-11.2	102	0.00
81	Benzo(a)anthracene	40.000	39.156	2.1	91	0.00
82	3,3'-Dichlorobenzidine	40.000	38.772	3.1	94	0.00
83	Chrysene	40.000	39.122	2.2	93	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.469	1.3	90	0.00
85 c	Di-n-octyl phthalate	40.000	39.874	0.3	93	0.00
86 I	Perylene-d12	20.000	20.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.490	1.3	95	0.00
88	Benzo(b)fluoranthene	40.000	40.864	-2.2	100	0.00
89	Benzo(k)fluoranthene	40.000	40.680	-1.7	102	0.00
90 C	Benzo(a)pyrene	40.000	40.528	-1.3	97	0.00
91	Dibenzo(a,h)anthracene	40.000	39.648	0.9	95	0.00
92	Benzo(g,h,i)perylene	40.000	42.521	-6.3	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031822\
Data File : BF127390.D
Acq On : 18 Mar 2022 14:20
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 19 01:51:15 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 19 01:50:04 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0