

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031824\
 Data File : BF137622.D
 Acq On : 18 Mar 2024 10:36
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 18 11:35:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 14 14:45:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	81	0.00
2	1,4-Dioxane	40.000	36.997	7.5	73	0.00
3	Pyridine	40.000	38.438	3.9	71	0.00
4	n-Nitrosodimethylamine	40.000	34.769	13.1	69	0.00
5 S	2-Fluorophenol	80.000	77.869	2.7	75	0.00
6	Aniline	40.000	36.579	8.6	70	0.00
7 S	Phenol-d6	80.000	72.906	8.9	73	0.00
8	2-Chlorophenol	40.000	37.795	5.5	75	0.00
9	Benzaldehyde	40.000	37.194	7.0	70	0.00
10 C	Phenol	40.000	36.865	7.8	72	0.00
11	bis(2-Chloroethyl)ether	40.000	36.038	9.9	70	0.00
12	1,3-Dichlorobenzene	40.000	37.662	5.8	75	0.00
13 C	1,4-Dichlorobenzene	40.000	37.571	6.1	75	0.00
14	1,2-Dichlorobenzene	40.000	37.493	6.3	76	0.00
15	Benzyl Alcohol	40.000	37.730	5.7	71	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	33.316	16.7	67	0.00
17	2-Methylphenol	40.000	37.239	6.9	74	0.00
18	Hexachloroethane	40.000	38.195	4.5	76	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	33.821	15.4	69	0.00
20	3+4-Methylphenols	40.000	37.066	7.3	74	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	77	0.00
22	Acetophenone	40.000	38.276	4.3	74	0.00
23 S	Nitrobenzene-d5	80.000	75.744	5.3	71	0.00
24	Nitrobenzene	40.000	37.976	5.1	72	0.00
25	Isophorone	40.000	36.190	9.5	69	0.00
26 C	2-Nitrophenol	40.000	40.444	-1.1	75	0.00
27	2,4-Dimethylphenol	40.000	38.794	3.0	73	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.766	5.6	71	0.00
29 C	2,4-Dichlorophenol	40.000	39.101	2.2	73	0.00
30	1,2,4-Trichlorobenzene	40.000	39.442	1.4	76	0.00
31	Naphthalene	40.000	37.932	5.2	74	0.00
32	Benzoic acid	40.000	32.345	19.1	64	0.00
33	4-Chloroaniline	40.000	38.665	3.3	70	0.00
34 C	Hexachlorobutadiene	40.000	39.308	1.7	75	0.00
35	Caprolactam	40.000	39.335	1.7	74	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.498	6.3	71	0.00
37	2-Methylnaphthalene	40.000	37.762	5.6	73	0.00
38	1-Methylnaphthalene	40.000	37.484	6.3	73	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	76	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.469	1.3	74	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.385	-1.0	81	0.00
42 S	2,4,6-Tribromophenol	80.000	78.326	2.1	74	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.318	1.7	72	0.00
44	2,4,5-Trichlorophenol	40.000	38.026	4.9	70	0.00
45 S	2-Fluorobiphenyl	80.000	74.039	7.5	72	0.00
46	1,1'-Biphenyl	40.000	38.233	4.4	72	0.00
47	2-Chloronaphthalene	40.000	38.318	4.2	72	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031824\
 Data File : BF137622.D
 Acq On : 18 Mar 2024 10:36
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 18 11:35:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 14 14:45:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.782	3.0	69	0.00
49	Acenaphthylene	40.000	37.635	5.9	71	0.00
50	Dimethylphthalate	40.000	37.259	6.9	71	0.00
51	2,6-Dinitrotoluene	40.000	39.309	1.7	72	0.00
52 C	Acenaphthene	40.000	37.099	7.3	71	0.00
53	3-Nitroaniline	40.000	39.499	1.3	70	0.00
54 P	2,4-Dinitrophenol	40.000	37.656	5.9	72	0.00
55	Dibenzofuran	40.000	36.958	7.6	70	0.00
56 P	4-Nitrophenol	40.000	38.192	4.5	72	0.00
57	2,4-Dinitrotoluene	40.000	40.011	-0.0	72	0.00
58	Fluorene	40.000	36.513	8.7	71	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.353	6.6	70	0.00
60	Diethylphthalate	40.000	38.164	4.6	73	0.00
61	4-Chlorophenyl-phenylether	40.000	36.933	7.7	73	0.00
62	4-Nitroaniline	40.000	37.393	6.5	69	0.00
63	Azobenzene	40.000	35.190	12.0	66	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	73	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.437	-1.1	70	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.583	3.5	71	0.00
67	4-Bromophenyl-phenylether	40.000	40.037	-0.1	72	0.00
68	Hexachlorobenzene	40.000	39.578	1.1	73	0.00
69	Atrazine	40.000	40.506	-1.3	74	0.00
70 C	Pentachlorophenol	40.000	36.270	9.3	63	0.00
71	Phenanthrene	40.000	37.450	6.4	70	0.00
72	Anthracene	40.000	37.994	5.0	71	0.00
73	Carbazole	40.000	38.537	3.7	71	0.00
74	Di-n-butylphthalate	40.000	39.347	1.6	72	0.00
75 C	Fluoranthene	40.000	38.264	4.3	72	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	87	0.00
77	Benzydine	40.000	26.268	34.3#	49	0.00
78	Pyrene	40.000	34.825	12.9	72	0.00
79 S	Terphenyl-d14	80.000	70.187	12.3	75	0.00
80	Butylbenzylphthalate	40.000	49.691	-24.2	103	0.00
81	Benzo(a)anthracene	40.000	37.177	7.1	80	0.00
82	3,3'-Dichlorobenzidine	40.000	42.207	-5.5	82	0.00
83	Chrysene	40.000	38.179	4.6	81	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	50.261	-25.7#	103	0.00
85 c	Di-n-octyl phthalate	40.000	41.992	-5.0	92	0.00
86 I	Perylene-d12	20.000	20.000	0.0	82	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	35.991	10.0	70	0.00
88	Benzo(b)fluoranthene	40.000	40.457	-1.1	82	0.00
89	Benzo(k)fluoranthene	40.000	35.855	10.4	76	0.00
90 C	Benzo(a)pyrene	40.000	38.039	4.9	75	0.00
91	Dibenzo(a,h)anthracene	40.000	35.281	11.8	68	0.00
92	Benzo(g,h,i)perylene	40.000	36.190	9.5	70	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031824\
Data File : BF137622.D
Acq On : 18 Mar 2024 10:36
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 18 11:35:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 14 14:45:20 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0