

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF032019\
 Data File : BF113185.D
 Acq On : 20 Mar 2019 21:56
 Operator : JU/SJ
 Sample : K1971-10 10X
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-15(6-7)-031519

Quant Time: Mar 21 15:40:02 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF022719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 21 12:13:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.72	152	116450	20.00	ng	0.00
21) Naphthalene-d8	8.00	136	464841	20.00	ng	0.00
39) Acenaphthene-d10	9.76	164	227093	20.00	ng	0.00
64) Phenanthrene-d10	11.23	188	426078	20.00	ng	0.00
76) Chrysene-d12	13.87	240	305872	20.00	ng	0.00
87) Perylene-d12	15.30	264	219391	20.00	ng	0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.33	112	100446	13.42	ng	0.00
7) Phenol-d6	6.36	99	130230	13.85	ng	-0.01
23) Nitrobenzene-d5	7.28	82	72300	9.31	ng	-0.01
42) 2,4,6-Tribromophenol	10.55	330	37826	15.69	ng	0.00
45) 2-Fluorobiphenyl	9.08	172	175167	9.13	ng	0.00
79) Terphenyl-d14	12.82	244	168896	10.70	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF032019\
Data File : BF113185.D
Acq On : 20 Mar 2019 21:56
Operator : JU/SJ
Sample : K1971-10 10X
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-15(6-7)-031519

Quant Time: Mar 21 15:40:02 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF022719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 21 12:13:22 2019
Response via : Initial Calibration

