

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032122\
 Data File : BF127424.D
 Acq On : 21 Mar 2022 23:55
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 22 05:06:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 22 02:16:33 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	111	0.00
2	1,4-Dioxane	40.000	44.807	-12.0	136	0.00
3	Pyridine	40.000	39.871	0.3	112	0.00
4	n-Nitrosodimethylamine	40.000	39.544	1.1	112	0.00
5 S	2-Fluorophenol	80.000	76.024	5.0	108	0.00
6	Aniline	40.000	38.891	2.8	109	0.00
7 S	Phenol-d6	80.000	76.602	4.2	109	0.00
8	2-Chlorophenol	40.000	38.726	3.2	109	0.00
9	Benzaldehyde	40.000	40.857	-2.1	111	0.00
10 C	Phenol	40.000	38.714	3.2	108	0.00
11	bis(2-Chloroethyl)ether	40.000	38.834	2.9	110	0.00
12	1,3-Dichlorobenzene	40.000	38.671	3.3	111	0.00
13 C	1,4-Dichlorobenzene	40.000	38.719	3.2	110	0.00
14	1,2-Dichlorobenzene	40.000	36.617	8.5	110	0.00
15	Benzyl Alcohol	40.000	38.659	3.4	108	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.961	2.6	110	0.00
17	2-Methylphenol	40.000	39.158	2.1	111	0.00
18	Hexachloroethane	40.000	38.182	4.5	109	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.523	6.2	109	0.00
20	3+4-Methylphenols	40.000	38.305	4.2	110	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	110	0.00
22	Acetophenone	40.000	37.496	6.3	108	0.00
23 S	Nitrobenzene-d5	80.000	77.304	3.4	110	0.00
24	Nitrobenzene	40.000	38.577	3.6	108	0.00
25	Isophorone	40.000	39.481	1.3	109	0.00
26 C	2-Nitrophenol	40.000	39.374	1.6	109	0.00
27	2,4-Dimethylphenol	40.000	39.047	2.4	107	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.237	1.9	111	0.00
29 C	2,4-Dichlorophenol	40.000	38.826	2.9	108	0.00
30	1,2,4-Trichlorobenzene	40.000	38.219	4.5	109	0.00
31	Naphthalene	40.000	38.468	3.8	109	0.00
32	Benzoic acid	40.000	38.434	3.9	107	0.00
33	4-Chloroaniline	40.000	40.119	-0.3	113	0.00
34 C	Hexachlorobutadiene	40.000	39.001	2.5	110	0.00
35	Caprolactam	40.000	39.652	0.9	106	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.214	2.0	110	0.00
37	2-Methylnaphthalene	40.000	38.790	3.0	109	0.00
38	1-Methylnaphthalene	40.000	38.092	4.8	109	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	107	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.149	4.6	107	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.000	2.5	109	0.00
42 S	2,4,6-Tribromophenol	80.000	75.728	5.3	106	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.115	-0.3	109	0.00
44	2,4,5-Trichlorophenol	40.000	41.244	-3.1	107	0.00
45 S	2-Fluorobiphenyl	80.000	74.963	6.3	107	0.00
46	1,1'-Biphenyl	40.000	38.005	5.0	108	0.00
47	2-Chloronaphthalene	40.000	38.480	3.8	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032122\
 Data File : BF127424.D
 Acq On : 21 Mar 2022 23:55
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 22 05:06:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 22 02:16:33 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.578	1.1	109	0.00
49	Acenaphthylene	40.000	38.278	4.3	106	0.00
50	Dimethylphthalate	40.000	38.839	2.9	108	0.00
51	2,6-Dinitrotoluene	40.000	38.951	2.6	108	0.00
52 C	Acenaphthene	40.000	38.413	4.0	106	0.00
53	3-Nitroaniline	40.000	38.662	3.3	106	0.00
54 P	2,4-Dinitrophenol	40.000	34.675	13.3	92	0.00
55	Dibenzofuran	40.000	40.670	-1.7	106	0.00
56 P	4-Nitrophenol	40.000	39.318	1.7	100	0.00
57	2,4-Dinitrotoluene	40.000	38.273	4.3	106	0.00
58	Fluorene	40.000	39.296	1.8	106	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.889	2.8	110	0.00
60	Diethylphthalate	40.000	37.380	6.5	106	0.00
61	4-Chlorophenyl-phenylether	40.000	36.027	9.9	107	0.00
62	4-Nitroaniline	40.000	38.073	4.8	106	0.00
63	Azobenzene	40.000	37.270	6.8	105	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	105	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.809	0.5	101	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.889	2.8	106	0.00
67	4-Bromophenyl-phenylether	40.000	37.914	5.2	102	0.00
68	Hexachlorobenzene	40.000	38.592	3.5	105	0.00
69	Atrazine	40.000	40.555	-1.4	106	0.00
70 C	Pentachlorophenol	40.000	38.413	4.0	101	0.00
71	Phenanthrene	40.000	37.960	5.1	105	0.00
72	Anthracene	40.000	38.025	4.9	104	0.00
73	Carbazole	40.000	37.792	5.5	103	0.00
74	Di-n-butylphthalate	40.000	39.077	2.3	105	0.00
75 C	Fluoranthene	40.000	37.644	5.9	102	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzydine	40.000	38.799	3.0	109	0.00
78	Pyrene	40.000	39.380	1.5	105	0.00
79 S	Terphenyl-d14	80.000	78.040	2.4	104	0.00
80	Butylbenzylphthalate	40.000	40.682	-1.7	102	0.00
81	Benzo(a)anthracene	40.000	38.377	4.1	100	0.00
82	3,3'-Dichlorobenzidine	40.000	38.479	3.8	101	0.00
83	Chrysene	40.000	38.292	4.3	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.395	-3.5	104	0.00
85 c	Di-n-octyl phthalate	40.000	40.638	-1.6	103	0.00
86 I	Perylene-d12	20.000	20.000	0.0	108	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.868	-7.2	109	0.00
88	Benzo(b)fluoranthene	40.000	39.352	1.6	105	0.00
89	Benzo(k)fluoranthene	40.000	34.444	13.9	101	0.00
90 C	Benzo(a)pyrene	40.000	38.743	3.1	106	0.00
91	Dibenzo(a,h)anthracene	40.000	42.954	-7.4	109	0.00
92	Benzo(g,h,i)perylene	40.000	42.370	-5.9	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032122\
Data File : BF127424.D
Acq On : 21 Mar 2022 23:55
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 22 05:06:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 22 02:16:33 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0