

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032122\
 Data File : BF127429.D
 Acq On : 22 Mar 2022 02:18
 Operator : CG\JU
 Sample : N1958-16MSD
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 C-1-0-1MSD

Quant Time: Mar 22 05:07:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 22 02:16:33 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Christian Giraldo 03/22/2022
 Supervised By : Jagrut Upadhyay 03/22/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.851	152	99322	20.000	ng	0.00
21) Naphthalene-d8	8.133	136	362456	20.000	ng	0.00
39) Acenaphthene-d10	9.892	164	210082	20.000	ng	0.00
64) Phenanthrene-d10	11.380	188	368090	20.000	ng	0.00
76) Chrysene-d12	14.027	240	231482	20.000	ng	0.00
86) Perylene-d12	15.504	264	178909	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	749882	120.515	ng	0.03
7) Phenol-d6	6.498	99	960557	112.935	ng	0.01
23) Nitrobenzene-d5	7.422	82	757542	89.340	ng	0.00
42) 2,4,6-Tribromophenol	10.686	330	300985	134.494	ng	0.00
45) 2-Fluorobiphenyl	9.210	172	1270149	88.285	ng	0.00
79) Terphenyl-d14	12.968	244	1393232	98.528	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.834	88	118178	43.243	ng	# 78
3) Pyridine	3.551	79	266047	34.491	ng	97
4) n-Nitrosodimethylamine	3.498	42	204654	48.324	ng	94
6) Aniline	6.522	93	216950	21.375	ng	# 88
8) 2-Chlorophenol	6.639	128	311051	48.918	ng	95
9) Benzaldehyde	6.410	77	139501	30.408	ng	98
10) Phenol	6.510	94	386019	41.674	ng	98
11) bis(2-Chloroethyl)ether	6.592	93	359510	53.043	ng	98
12) 1,3-Dichlorobenzene	6.792	146	308830	43.182	ng	96
13) 1,4-Dichlorobenzene	6.869	146	317266	44.102	ng	100
14) 1,2-Dichlorobenzene	7.016	146	301676	43.056	ng	97
15) Benzyl Alcohol	7.004	79	270897	43.825	ng	97
16) 2,2'-oxybis(1-Chloropr...	7.122	45	566403	46.440	ng	80
17) 2-Methylphenol	7.110	107	275675	51.372	ng	94
18) Hexachloroethane	7.357	117	117093	44.460	ng	97
19) n-Nitroso-di-n-propyla...	7.269	70	243579	47.776	ng	98
20) 3+4-Methylphenols	7.263	107	306537	46.313	ng	92
22) Acetophenone	7.263	105	429683	46.376	ng	97
24) Nitrobenzene	7.439	77	394517	48.787	ng	99
25) Isophorone	7.675	82	717322	53.115	ng	99
26) 2-Nitrophenol	7.751	139	163626	48.246	ng	99
27) 2,4-Dimethylphenol	7.792	122	287193	56.294	ng	99
28) bis(2-Chloroethoxy)met...	7.880	93	418504	47.704	ng	100
29) 2,4-Dichlorophenol	7.998	162	280172	48.376	ng	98
30) 1,2,4-Trichlorobenzene	8.075	180	301026	44.605	ng	96
31) Naphthalene	8.157	128	890788	46.291	ng	98
32) Benzoic acid	7.957	122	152632	51.546	ng	93
33) 4-Chloroaniline	8.210	127	74055	9.931	ng	98
34) Hexachlorobutadiene	8.263	225	188564	44.000	ng	97
35) Caprolactam	8.598	113	67389m	37.743	ng	
36) 4-Chloro-3-methylphenol	8.704	107	310522	49.417	ng	99
37) 2-Methylnaphthalene	8.845	142	600338	48.518	ng	100
38) 1-Methylnaphthalene	8.945	142	565531	46.221	ng	99
40) 1,2,4,5-Tetrachloroben...	9.010	216	310286	46.452	ng	99
41) Hexachlorocyclopentadiene	8.992	237	259611	99.951	ng	97
43) 2,4,6-Trichlorophenol	9.127	196	192833	47.758	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032122\
 Data File : BF127429.D
 Acq On : 22 Mar 2022 02:18
 Operator : CG\JU
 Sample : N1958-16MSD
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 C-1-0-1MSD

Quant Time: Mar 22 05:07:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 22 02:16:33 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 03/22/2022
 Supervised By : Jagrut Upadhyay 03/22/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.180	196	213387	51.340	ng	96
46) 1,1'-Biphenyl	9.310	154	737859	46.504	ng	99
47) 2-Chloronaphthalene	9.339	162	581894	46.572	ng	99
48) 2-Nitroaniline	9.445	65	226022	47.200	ng	97
49) Acenaphthylene	9.757	152	919423	48.675	ng	99
50) Dimethylphthalate	9.616	163	704991	48.237	ng	99
51) 2,6-Dinitrotoluene	9.686	165	153167	49.926	ng	95
52) Acenaphthene	9.927	154	530475	48.133	ng	99
53) 3-Nitroaniline	9.857	138	89714	26.659	ng	95
54) 2,4-Dinitrophenol	9.974	184	163617	102.880	ng	99
55) Dibenzofuran	10.098	168	773347	48.920	ng	99
56) 4-Nitrophenol	10.033	139	159997	93.411	ng	99
57) 2,4-Dinitrotoluene	10.092	165	191256	48.433	ng	94
58) Fluorene	10.445	166	628372	49.458	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.222	232	184484	50.203	ng	95
60) Diethylphthalate	10.316	149	618274	46.317	ng	99
61) 4-Chlorophenyl-phenyle...	10.433	204	331155	43.841	ng	95
62) 4-Nitroaniline	10.480	138	136339	43.252	ng	98
63) Azobenzene	10.592	77	713351	45.289	ng	99
65) 4,6-Dinitro-2-methylph...	10.504	198	112653	49.738	ng	98
66) n-Nitrosodiphenylamine	10.557	169	536219	47.856	ng	99
67) 4-Bromophenyl-phenylether	10.921	248	205698	47.864	ng	97
68) Hexachlorobenzene	10.986	284	230172	49.067	ng	95
69) Atrazine	11.080	200	191059	51.219	ng	97
70) Pentachlorophenol	11.192	266	206250	111.022	ng	97
71) Phenanthrene	11.410	178	906196	47.222	ng	100
72) Anthracene	11.463	178	919162	48.266	ng	99
73) Carbazole	11.621	167	824870	45.293	ng	99
74) Di-n-butylphthalate	11.933	149	923280	48.977	ng	99
75) Fluoranthene	12.598	202	1002018	47.473	ng	99
77) Benzidine	12.721	184	137736	24.955	ng	96
78) Pyrene	12.827	202	1002276	50.451	ng	100
80) Butylbenzylphthalate	13.433	149	345314	52.813	ng	100
81) Benzo(a)anthracene	14.015	228	761068	49.164	ng	98
82) 3,3'-Dichlorobenzidine	13.974	252	130779	28.745	ng	100
83) Chrysene	14.057	228	702092	47.895	ng	100
84) Bis(2-ethylhexyl)phtha...	13.986	149	416780	55.143	ng	99
85) Di-n-octyl phthalate	14.598	149	541001	53.789	ng	100
87) Indeno(1,2,3-cd)pyrene	17.009	276	633568	52.335	ng	100
88) Benzo(b)fluoranthene	15.074	252	583640	51.738	ng	98
89) Benzo(k)fluoranthene	15.104	252	551386	47.150	ng	99
90) Benzo(a)pyrene	15.445	252	549362	59.340	ng	99
91) Dibenzo(a,h)anthracene	17.015	278	561856	56.215	ng	97
92) Benzo(g,h,i)perylene	17.462	276	559789	54.186	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.851 min Scan# 71
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD



Tgt Ion:152 Resp: 9932
Ion Ratio Lower Upper
152 100
150 139.0 117.2 175.8
115 58.2 49.8 74.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022



#2
1,4-Dioxane
Concen: 43.243 ng
RT: 2.834 min Scan# 93
Delta R.T. 0.229 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18


Tgt Ion: 88 Resp: 118178
Ion Ratio Lower Upper
88 100
58 91.7 77.0 115.4
43 0.0 31.3 46.9





#3
Pyridine
Concen: 34.491 ng
RT: 3.551 min Scan# 215
Delta R.T. 0.182 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion: 79 Resp: 26604
Ion Ratio Lower Upper
79 100
52 89.9 74.2 111.4
51 43.3 36.7 55.1

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#4
n-Nitrosodimethylamine
Concen: 48.324 ng
RT: 3.498 min Scan# 206
Delta R.T. 0.153 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18



Tgt Ion: 42 Resp: 204654
Ion Ratio Lower Upper
42 100
74 108.6 92.4 138.6
44 7.6 6.0 9.0









#5
2-Fluorophenol
Concen: 120.515 ng
RT: 5.498 min Scan# 541
Delta R.T. 0.029 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18

Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD

Tgt Ion:112 Resp: 74988
Ion Ratio Lower Upper
112 100
64 73.0 58.4 87.6
63 38.2 30.1 45.1




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022



#6
Aniline
Concen: 21.375 ng
RT: 6.522 min Scan# 720
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18


Tgt Ion: 93 Resp: 216950
Ion Ratio Lower Upper
93 100
66 53.3 36.2 54.4
65 29.3 18.4 27.6#





#7
Phenol-d6
Concen: 112.935 ng
RT: 6.498 min Scan# 714
Delta R.T. 0.012 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion: 99 Resp: 96055
Ion Ratio Lower Upper
99 100
42 21.9 17.8 26.6
71 37.2 29.6 44.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#8
2-Chlorophenol
Concen: 48.918 ng
RT: 6.639 min Scan# 740
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:128 Resp: 311051
Ion Ratio Lower Upper
128 100
130 31.7 11.3 51.3
64 63.3 38.4 78.4





#9
Benzaldehyde
Concen: 30.408 ng
RT: 6.410 min Scan# 700
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion: 77 Resp: 139500
Ion Ratio Lower Upper
77 100
105 90.4 66.8 106.8
106 84.4 64.3 104.3

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#10
Phenol
Concen: 41.674 ng
RT: 6.510 min Scan# 718
Delta R.T. 0.012 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18



Tgt Ion: 94 Resp: 386019
Ion Ratio Lower Upper
94 100
65 33.7 13.1 53.1
66 46.5 24.9 64.9









#11
bis(2-Chloroethyl)ether
Concen: 53.043 ng
RT: 6.592 min Scan# 711
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18

Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD

Tgt Ion: 93 Resp: 359510
Ion Ratio Lower Upper
93 100
63 87.7 69.6 109.6
95 31.8 11.6 51.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022



#12
1,3-Dichlorobenzene
Concen: 43.182 ng
RT: 6.792 min Scan# 766
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18


Tgt Ion:146 Resp: 308830
Ion Ratio Lower Upper
146 100
148 64.0 50.6 76.0
75 34.3 32.4 48.6





#13
1,4-Dichlorobenzene
Concen: 44.102 ng
RT: 6.869 min Scan# 71
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:146 Resp: 317260
Ion Ratio Lower Upper
146 100
148 63.7 51.3 76.9
111 42.9 34.2 51.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#14
1,2-Dichlorobenzene
Concen: 43.056 ng
RT: 7.016 min Scan# 804
Delta R.T. -0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:146 Resp: 301676
Ion Ratio Lower Upper
146 100
148 62.5 51.4 77.2
111 48.7 36.4 54.6





#15
Benzyl Alcohol
Concen: 43.825 ng
RT: 7.004 min Scan# 801
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD



Tgt Ion: 79 Resp: 270897
Ion Ratio Lower Upper
79 100
108 70.0 52.2 78.2
77 60.4 48.3 72.5




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022



#16
2,2'-oxybis(1-Chloropropane)
Concen: 46.440 ng
RT: 7.122 min Scan# 822
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18


Tgt Ion: 45 Resp: 566403
Ion Ratio Lower Upper
45 100
77 20.5 0.0 32.9
79 17.7 0.0 29.9





#17
2-Methylphenol
Concen: 51.372 ng
RT: 7.110 min Scan# 819
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD



Tgt Ion:107 Resp: 27567
Ion Ratio Lower Upper
107 100
108 115.2 90.7 136.1
77 61.6 42.9 64.3
79 60.4 43.0 64.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022



#18
Hexachloroethane
Concen: 44.460 ng
RT: 7.357 min Scan# 862
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18

Tgt Ion:117 Resp: 117093
Ion Ratio Lower Upper
117 100
119 97.6 77.6 116.4
201 115.9 88.2 132.2






#19
n-Nitroso-di-n-propylamine
Concen: 47.776 ng
RT: 7.269 min Scan# 846
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18

Instrument :

BNA_F

ClientSampleId :

C-1-0-1MSD






Tgt Ion: 70 Resp: 243579
Ion Ratio Lower Upper
70 100
42 67.7 53.0 79.6
101 8.8 6.8 10.2
130 17.7 13.8 20.6


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#20
3+4-Methylphenols
Concen: 46.313 ng
RT: 7.263 min Scan# 846
Delta R.T. -0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:107 Resp: 306537
Ion Ratio Lower Upper
107 100
108 86.0 70.7 110.7
77 160.0 127.9 167.9
79 32.5 16.8 56.8





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.133 min Scan# 994
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:136 Resp: 362450
Ion Ratio Lower Upper
136 100
137 10.8 8.6 13.0
54 10.9 9.2 13.8
68 6.8 5.4 8.0

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#22
Acetophenone
Concen: 46.376 ng
RT: 7.263 min Scan# 846
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:105 Resp: 429683
Ion Ratio Lower Upper
105 100
71 8.2 7.3 10.9
51 46.1 38.8 58.2
120 21.1 16.5 24.7





#23
Nitrobenzene-d5
Concen: 89.340 ng
RT: 7.422 min Scan# 81
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion: 82 Resp: 75754
Ion Ratio Lower Upper
82 100
128 37.4 30.6 45.8
54 61.1 48.2 72.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#24
Nitrobenzene
Concen: 48.787 ng
RT: 7.439 min Scan# 876
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion: 77 Resp: 394517
Ion Ratio Lower Upper
77 100
123 38.8 32.0 48.0
65 15.2 12.2 18.2





#25
Isophorone
Concen: 53.115 ng
RT: 7.675 min Scan# 916
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
Client SampleId :
C-1-0-1MSD




Tgt Ion: 82 Resp: 71732
Ion Ratio Lower Upper
82 100
95 7.5 5.8 8.8
138 14.2 11.6 17.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#26
2-Nitrophenol
Concen: 48.246 ng
RT: 7.751 min Scan# 929
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18



Tgt Ion:139 Resp: 163626
Ion Ratio Lower Upper
139 100
109 35.7 28.2 42.2
65 70.1 55.3 82.9






#27
2,4-Dimethylphenol
Concen: 56.294 ng
RT: 7.792 min Scan# 91
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18

Instrument :

BNA_F

ClientSampleId :

C-1-0-1MSD






Tgt Ion:122 Resp: 28719
Ion Ratio Lower Upper
122 100
107 116.9 93.0 139.4
121 57.5 44.9 67.3


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#28
bis(2-Chloroethoxy)methane
Concen: 47.704 ng
RT: 7.880 min Scan# 951
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion: 93 Resp: 418504
Ion Ratio Lower Upper
93 100
95 32.8 26.3 39.5
123 9.7 8.0 12.0





#29
2,4-Dichlorophenol
Concen: 48.376 ng
RT: 7.998 min Scan# 91
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:162 Resp: 280172
Ion Ratio Lower Upper
162 100
164 63.7 43.7 83.7
98 40.0 23.5 63.5

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#30
1,2,4-Trichlorobenzene
Concen: 44.605 ng
RT: 8.075 min Scan# 984
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:180 Resp: 301026
Ion Ratio Lower Upper
180 100
182 97.4 73.8 110.6
145 29.2 23.3 34.9





#31
Naphthalene
Concen: 46.291 ng
RT: 8.157 min Scan# 998
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:128 Resp: 890788
Ion Ratio Lower Upper
128 100
129 11.0 8.4 12.6
127 13.4 10.2 15.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#32
Benzoic acid
Concen: 51.546 ng
RT: 7.957 min Scan# 964
Delta R.T. 0.024 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:122 Resp: 152632
Ion Ratio Lower Upper
122 100
105 128.7 117.1 157.1
77 105.1 93.0 133.0





#33
4-Chloroaniline
Concen: 9.931 ng
RT: 8.210 min Scan# 1007
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:127 Resp: 7405
Ion Ratio Lower Upper
127 100
129 30.7 25.9 38.9
65 39.0 31.4 47.0
92 20.9 18.3 27.5

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#34
Hexachlorobutadiene
Concen: 44.000 ng
RT: 8.263 min Scan# 1016
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:225 Resp: 188564
Ion Ratio Lower Upper
225 100
223 63.9 50.6 76.0
227 66.1 49.6 74.4





#35
Caprolactam
Concen: 37.743 ng m
RT: 8.598 min Scan# 1073
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:113 Resp: 67389
Ion Ratio Lower Upper
113 100
55 233.0 218.7 258.7
56 171.7 156.2 196.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#36
4-Chloro-3-methylphenol
Concen: 49.417 ng
RT: 8.704 min Scan# 1091
Delta R.T. 0.012 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18



Tgt Ion:107 Resp: 310522
Ion Ratio Lower Upper
107 100
144 23.7 19.3 28.9
142 74.1 59.8 89.8





#37

2-Methylnaphthalene

Concen: 48.518 ng

RT: 8.845 min Scan# 1115

Delta R.T. 0.000 min

Lab File: BF127429.D

Acq: 22 Mar 2022 02:18

Instrument :

BNA_F

ClientSampleId :

C-1-0-1MSD

Tgt Ion:142 Resp: 600338

Ion Ratio Lower Upper

142 100





141 88.7 70.6 106.0

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 03/22/2022

Supervised By :Jagrut Upadhyay 03/22/2022



#38

1-Methylnaphthalene

Concen: 46.221 ng

RT: 8.945 min Scan# 1132

Delta R.T. 0.000 min

Lab File: BF127429.D




Acq: 22 Mar 2022 02:18


Tgt Ion:142 Resp: 565531

Ion Ratio Lower Upper

142 100


141 91.9 74.2 111.2





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.892 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:164 Resp: 21008
Ion Ratio Lower Upper
164 100
162 99.6 81.2 121.8
160 43.1 34.6 52.0

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#40
1,2,4,5-Tetrachlorobenzene
Concen: 46.452 ng
RT: 9.010 min Scan# 1143
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:216 Resp: 310286
Ion Ratio Lower Upper
216 100
214 79.2 63.0 94.4
179 20.1 15.5 23.3
108 19.4 13.8 20.6





#41
Hexachlorocyclopentadiene
Concen: 99.951 ng
RT: 8.992 min Scan# 1140
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:237 Resp: 25961
Ion Ratio Lower Upper
237 100
235 63.2 40.4 80.4
272 12.0 0.0 31.8

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#42
2,4,6-Tribromophenol
Concen: 134.494 ng
RT: 10.686 min Scan# 1428
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:330 Resp: 300985
Ion Ratio Lower Upper
330 100
332 94.2 79.1 118.7
141 31.0 26.2 39.4





#43
2,4,6-Trichlorophenol
Concen: 47.758 ng
RT: 9.127 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:196 Resp: 19283
Ion Ratio Lower Upper
196 100
198 94.0 77.8 116.6
200 32.0 25.9 38.9

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#44
2,4,5-Trichlorophenol
Concen: 51.340 ng
RT: 9.180 min Scan# 1172
Delta R.T. 0.012 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:196 Resp: 213387
Ion Ratio Lower Upper
196 100
198 95.4 75.3 112.9
97 52.3 46.8 70.2
132 25.5 22.7 34.1





#45
2-Fluorobiphenyl
Concen: 88.285 ng
RT: 9.210 min Scan# 1177
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:172 Resp: 1270149
Ion Ratio Lower Upper
172 100
171 35.1 28.4 42.6
170 23.3 18.6 27.8

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#46
1,1'-Biphenyl
Concen: 46.504 ng
RT: 9.310 min Scan# 1194
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18



Tgt Ion:154 Resp: 737859
Ion Ratio Lower Upper
154 100
153 39.3 20.3 60.3
76 13.9 0.0 33.9





#47

2-Chloronaphthalene

Concen: 46.572 ng

RT: 9.339 min Scan# 1199

Delta R.T. 0.000 min

Lab File: BF127429.D

Acq: 22 Mar 2022 02:18

Instrument :

BNA_F

ClientSampleId :





C-1-0-1MSD

Tgt Ion:	162	Resp:	581894
Ion Ratio	Lower	Upper	
162	100		
127	40.8	31.8	47.6
164	32.9	26.4	39.6

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022



#48

2-Nitroaniline

Concen: 47.200 ng




RT: 9.445 min Scan# 1217


Delta R.T. 0.006 min

Lab File: BF127429.D

Acq: 22 Mar 2022 02:18


Tgt Ion:	65	Resp:	226022
Ion Ratio	Lower	Upper	
65	100		
92	63.8	50.0	75.0
138	83.8	64.1	96.1





#49
Acenaphthylene
Concen: 48.675 ng
RT: 9.757 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:152 Resp: 91942
Ion Ratio Lower Upper
152 100
151 20.2 15.8 23.8
153 13.3 10.5 15.7

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#50
Dimethylphthalate
Concen: 48.237 ng
RT: 9.616 min Scan# 1246
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:163 Resp: 704991
Ion Ratio Lower Upper
163 100
194 3.7 3.3 4.9
164 10.4 7.8 11.8





#51
2,6-Dinitrotoluene
Concen: 49.926 ng
RT: 9.686 min Scan# 1257
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:165 Resp: 15316
Ion Ratio Lower Upper
165 100
63 67.5 58.3 87.5
89 62.5 51.5 77.3

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#52
Acenaphthene
Concen: 48.133 ng
RT: 9.927 min Scan# 1299
Delta R.T. 0.005 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:154 Resp: 530475
Ion Ratio Lower Upper
154 100
153 111.1 87.6 131.4
152 53.8 42.9 64.3





#53
3-Nitroaniline
Concen: 26.659 ng
RT: 9.857 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:138 Resp: 89714
Ion Ratio Lower Upper
138 100
108 12.2 9.4 14.0
92 138.5 106.1 159.1

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#54
2,4-Dinitrophenol
Concen: 102.880 ng
RT: 9.974 min Scan# 1307
Delta R.T. 0.005 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:184 Resp: 163617
Ion Ratio Lower Upper
184 100
63 72.1 58.1 87.1
154 64.8 53.2 79.8





#55
Dibenzenofuran
Concen: 48.920 ng
RT: 10.098 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:168 Resp: 77334
Ion Ratio Lower Upper
168 100
139 38.7 30.3 45.5
169 13.4 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#56
4-Nitrophenol
Concen: 93.411 ng
RT: 10.033 min Scan# 1317
Delta R.T. 0.011 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:139 Resp: 159997
Ion Ratio Lower Upper
139 100
109 87.1 64.3 104.3
65 137.0 117.5 157.5





#57
2,4-Dinitrotoluene
Concen: 48.433 ng
RT: 10.092 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:165 Resp: 191250
Ion Ratio Lower Upper
165 100
63 72.7 53.1 79.7
89 90.9 69.2 103.8
182 2.7 2.2 3.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#58
Fluorene
Concen: 49.458 ng
RT: 10.445 min Scan# 1387
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:166 Resp: 628372
Ion Ratio Lower Upper
166 100
165 96.7 78.1 117.1
167 12.9 10.5 15.7





#59
2,3,4,6-Tetrachlorophenol
Concen: 50.203 ng
RT: 10.222 min Scan# 1349
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:232 Resp: 184484
Ion Ratio Lower Upper
232 100
131 38.3 33.4 50.0
130 1.8 1.8 2.6
166 26.2 22.6 34.0

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#60
Diethylphthalate
Concen: 46.317 ng
RT: 10.316 min Scan# 1365
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:149 Resp: 618274
Ion Ratio Lower Upper
149 100
177 23.2 18.9 28.3
150 12.7 10.0 15.0





#61
4-Chlorophenyl-phenylether
Concen: 43.841 ng
RT: 10.433 min Scan# 1385
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:204 Resp: 33115
Ion Ratio Lower Upper
204 100
206 32.5 25.9 38.9
141 54.3 47.8 71.8

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#62
4-Nitroaniline
Concen: 43.252 ng
RT: 10.480 min Scan# 1393
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:138 Resp: 136339
Ion Ratio Lower Upper
138 100
92 62.6 40.5 80.5
108 87.3 69.2 109.2





#63
Azobenzene
Concen: 45.289 ng
RT: 10.592 min Scan# 1412
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion: 77 Resp: 71335
Ion Ratio Lower Upper
77 100
182 22.9 4.5 44.5
105 19.6 0.3 40.3
51 35.2 15.0 55.0

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.380 min Scan# 1546
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:188 Resp: 368090
Ion Ratio Lower Upper
188 100
94 9.6 7.4 11.0
80 11.3 8.0 12.0





#65
4,6-Dinitro-2-methylphenol
Concen: 49.738 ng
RT: 10.504 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:198 Resp: 11265
Ion Ratio Lower Upper
198 100
51 61.8 42.7 82.7
105 43.5 25.4 65.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#66
n-Nitrosodiphenylamine
Concen: 47.856 ng
RT: 10.557 min Scan# 1406
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:169 Resp: 536219
Ion Ratio Lower Upper
169 100
168 67.1 52.7 79.1
167 36.5 28.9 43.3





#67
4-Bromophenyl-phenylether
Concen: 47.864 ng
RT: 10.921 min Scan# 1468
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:248 Resp: 205698
Ion Ratio Lower Upper
248 100
250 95.8 76.6 115.0
141 71.5 52.2 78.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#68
Hexachlorobenzene
Concen: 49.067 ng
RT: 10.986 min Scan# 1479
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:284 Resp: 230172
Ion Ratio Lower Upper
284 100
142 32.7 23.8 35.8
249 32.1 23.9 35.9





#69
Atrazine
Concen: 51.219 ng
RT: 11.080 min Scan# 1495
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:200 Resp: 191059
Ion Ratio Lower Upper
200 100
173 25.3 2.9 42.9
215 47.0 28.0 68.0

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#70
Pentachlorophenol
Concen: 111.022 ng
RT: 11.192 min Scan# 1514
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:266 Resp: 206250
Ion Ratio Lower Upper
266 100
268 63.2 50.4 75.6
264 64.6 48.5 72.7





#71
Phenanthrene
Concen: 47.222 ng
RT: 11.410 min Scan# 1550
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:178 Resp: 906190
Ion Ratio Lower Upper
178 100
176 19.5 15.4 23.2
179 15.3 12.4 18.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#72
Anthracene
Concen: 48.266 ng
RT: 11.463 min Scan# 1560
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18



Tgt Ion:178 Resp: 919162
Ion Ratio Lower Upper
178 100
176 18.9 14.6 22.0
179 15.1 12.4 18.6





#73
 Carbazole
 Concen: 45.293 ng
 RT: 11.621 min Scan# 1587
 Delta R.T. 0.006 min
 Lab File: BF127429.D
 Acq: 22 Mar 2022 02:18

Instrument :

BNA_F

ClientSampleId :





C-1-0-1MSD

Tgt Ion:167 Resp: 824870
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.0 25.4
 139 12.2 10.5 15.7

Manual Integrations




APPROVED


Reviewed By :Christian Giraldo 03/22/2022
 Supervised By :Jagrut Upadhyay 03/22/2022



#74
 Di-n-butylphthalate
 Concen: 48.977 ng
 RT: 11.933 min Scan# 1640
 Delta R.T. 0.000 min
 Lab File: BF127429.D
 Acq: 22 Mar 2022 02:18

Tgt Ion:149 Resp: 923280
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.5 11.3
 104 5.5 4.2 6.2









#75
Fluoranthene
Concen: 47.473 ng
RT: 12.598 min Scan# 1753
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18

Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD

Tgt Ion:202 Resp: 1002018
Ion Ratio Lower Upper
202 100
101 10.0 0.0 29.0
203 17.8 0.0 37.7




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.027 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18


Tgt Ion:240 Resp: 231482
Ion Ratio Lower Upper
240 100
120 9.5 7.1 10.7
236 26.0 20.7 31.1





#77
Benzidine
Concen: 24.955 ng
RT: 12.721 min Scan# 1774
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:184 Resp: 137730
Ion Ratio Lower Upper
184 100
185 12.7 12.2 18.2
183 11.0 9.4 14.0

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#78
Pyrene
Concen: 50.451 ng
RT: 12.827 min Scan# 1792
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18



Tgt Ion:202 Resp: 1002276
Ion Ratio Lower Upper
202 100
200 21.4 17.1 25.7
203 17.7 14.3 21.5






#79
 Terphenyl-d14
 Concen: 98.528 ng
 RT: 12.968 min Scan# 1816
 Delta R.T. 0.000 min
 Lab File: BF127429.D
 Acq: 22 Mar 2022 02:18

Instrument :

BNA_F

ClientSampleId :

C-1-0-1MSD






Tgt Ion:244 Resp: 139323
 Ion Ratio Lower Upper
 244 100
 212 6.8 5.7 8.5
 122 8.5 6.2 9.2


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 03/22/2022
 Supervised By :Jagrut Upadhyay 03/22/2022




#80
 Butylbenzylphthalate
 Concen: 52.813 ng
 RT: 13.433 min Scan# 1895
 Delta R.T. 0.000 min
 Lab File: BF127429.D
 Acq: 22 Mar 2022 02:18



Tgt Ion:149 Resp: 345314
 Ion Ratio Lower Upper
 149 100
 91 86.3 69.1 103.7
 206 22.5 17.7 26.5






#81
Benzo(a)anthracene
Concen: 49.164 ng
RT: 14.015 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18

Instrument :

BNA_F

ClientSampleId :

C-1-0-1MSD






Tgt Ion:228 Resp: 761068
Ion Ratio Lower Upper
228 100
226 27.9 21.6 32.4
229 20.1 15.4 23.2


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#82
3,3'-Dichlorobenzidine
Concen: 28.745 ng
RT: 13.974 min Scan# 1987
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:252 Resp: 130779
Ion Ratio Lower Upper
252 100
254 62.5 50.2 75.4
126 11.0 8.7 13.1





#83
Chrysene
Concen: 47.895 ng
RT: 14.057 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:228 Resp: 702092
Ion Ratio Lower Upper
228 100
226 30.2 24.2 36.4
229 19.7 15.4 23.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#84
Bis(2-ethylhexyl)phthalate
Concen: 55.143 ng
RT: 13.986 min Scan# 1989
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:149 Resp: 416780
Ion Ratio Lower Upper
149 100
167 28.2 22.9 34.3
279 4.6 3.2 4.8





#85
Di-n-octyl phthalate
Concen: 53.789 ng
RT: 14.598 min Scan# 2093
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
Client SampleId :
C-1-0-1MSD




Tgt Ion:149 Resp: 54100
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 13.8 11.0 16.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.504 min Scan# 2247
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:264 Resp: 178909
Ion Ratio Lower Upper
264 100
260 24.2 19.1 28.7
265 21.2 17.4 26.2





#87
Indeno(1,2,3-cd)pyrene
Concen: 52.335 ng
RT: 17.009 min Scan# 211
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:276 Resp: 63356
Ion Ratio Lower Upper
276 100
138 21.7 17.1 25.7
277 25.1 19.9 29.9

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#88
Benzo(b)fluoranthene
Concen: 51.738 ng
RT: 15.074 min Scan# 2174
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:252 Resp: 583640
Ion Ratio Lower Upper
252 100
253 21.7 17.9 26.9
125 7.9 7.0 10.4





#89
Benzo(k)fluoranthene
Concen: 47.150 ng
RT: 15.104 min Scan# 2179
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:252 Resp: 551380
Ion Ratio Lower Upper
252 100
253 21.9 17.3 25.9
125 8.1 6.9 10.3

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022




#90
Benzo(a)pyrene
Concen: 59.340 ng
RT: 15.445 min Scan# 2237
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Tgt Ion:252 Resp: 549362
Ion Ratio Lower Upper
252 100
253 21.3 17.0 25.6
125 9.5 6.7 10.1





#91
Dibenzo(a,h)anthracene
Concen: 56.215 ng
RT: 17.015 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18




Instrument :
BNA_F
ClientSampleId :
C-1-0-1MSD




Tgt Ion:278 Resp: 561850
Ion Ratio Lower Upper
278 100
139 14.0 10.8 16.2
279 23.8 17.7 26.5

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 03/22/2022
Supervised By :Jagrut Upadhyay 03/22/2022



#92
Benzo(g,h,i)perylene
Concen: 54.186 ng
RT: 17.462 min Scan# 2580
Delta R.T. 0.006 min
Lab File: BF127429.D
Acq: 22 Mar 2022 02:18



Tgt Ion:276 Resp: 559789
Ion Ratio Lower Upper
276 100
277 23.6 19.0 28.4
138 18.7 14.6 22.0

