



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032122\
 Data File : BF127413.D
 Acq On : 21 Mar 2022 17:11
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Mar 22 02:16:33 2022






AutoFind: Scans 1586, 1587, 1588; Background Corrected with Scan 1580

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.9	45586	PASS
68	69	0.00	2	1.9	737	PASS
69	198	0.00	100	36.8	38209	PASS
70	69	0.00	2	0.5	202	PASS
127	198	10	80	43.9	45549	PASS
197	198	0.00	2	0.1	129	PASS
198	198	100	100	100.0	103861	PASS
199	198	5	9	6.8	7028	PASS
275	198	10	60	28.5	29592	PASS
365	198	1	100	3.6	3734	PASS
441	198	0.01	100	15.9	16545	PASS
442	442	50	100	100.0	106112	PASS
443	442	15	24	17.9	18981	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032122\
 Data File : BF127413.D
 Acq On : 21 Mar 2022 17:11
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Mar 22 03:17:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 22 02:16:33 2022
 Response via : Initial Calibration



TIC: BF127413.D\data.ms

(70) Pentachlorophenol (C)

11.186min (+ 0.000) 29510.69 ng




response 67535

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	61.53
264.00	60.60	59.65
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032122\
Data File : BF127413.D
Acq On : 21 Mar 2022 17:11
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 22 03:17:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 22 02:16:33 2022
Response via : Initial Calibration



TIC: BF127413.D\data.ms

(77) Benzidine

12.721min (0.000) 0.00 ng

response 586347

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.20	14.71
183.00	11.70	11.48
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/21/2022	BNA_F	<u>BF127413.D</u>
Compound Name	Response	Retention Time
DDT	285966	13.533
DDD	4449	13.239
DDE	425	12.898
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
4874	290840	1.68

Instrument :
BNA_F
ClientSampleId :
DFTPP