

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032522\
 Data File : BF127533.D
 Acq On : 25 Mar 2022 18:45
 Operator : CG\JU
 Sample : N2077-16
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-14A-3-3

Quant Time: Mar 26 02:21:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 22 02:16:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.845	152	85173	20.000	ng	0.00
21) Naphthalene-d8	8.127	136	313879	20.000	ng	0.00
39) Acenaphthene-d10	9.880	164	183242	20.000	ng	-0.01
64) Phenanthrene-d10	11.374	188	322367	20.000	ng	0.00
76) Chrysene-d12	14.027	240	199521	20.000	ng	0.00
86) Perylene-d12	15.509	264	171011	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.486	112	690154	129.341	ng	0.02
7) Phenol-d6	6.492	99	870180	119.304	ng	0.00
23) Nitrobenzene-d5	7.416	82	671107	91.395	ng	0.00
42) 2,4,6-Tribromophenol	10.680	330	258729	132.546	ng	0.00
45) 2-Fluorobiphenyl	9.204	172	1124597	89.618	ng	0.00
79) Terphenyl-d14	12.962	244	1260194	103.396	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032522\
Data File : BF127533.D
Acq On : 25 Mar 2022 18:45
Operator : CG\JU
Sample : N2077-16
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-14A-3-3

Quant Time: Mar 26 02:21:35 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 22 02:16:33 2022
Response via : Initial Calibration

