

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032522\
 Data File : BF127518.D
 Acq On : 25 Mar 2022 09:48
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 28 05:16:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 22 02:16:33 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	108	0.00
2	1,4-Dioxane	40.000	48.409	-21.0	142	0.00
3	Pyridine	40.000	43.787	-9.5	119	0.00
4	n-Nitrosodimethylamine	40.000	43.037	-7.6	118	0.00
5 S	2-Fluorophenol	80.000	86.418	-8.0	119	0.00
6	Aniline	40.000	43.581	-9.0	118	0.00
7 S	Phenol-d6	80.000	85.911	-7.4	119	0.00
8	2-Chlorophenol	40.000	43.993	-10.0	120	0.00
9	Benzaldehyde	40.000	42.725	-6.8	112	0.00
10 C	Phenol	40.000	43.172	-7.9	117	0.00
11	bis(2-Chloroethyl)ether	40.000	44.188	-10.5	121	0.00
12	1,3-Dichlorobenzene	40.000	42.353	-5.9	117	0.00
13 C	1,4-Dichlorobenzene	40.000	42.507	-6.3	117	0.00
14	1,2-Dichlorobenzene	40.000	40.085	-0.2	117	0.00
15	Benzyl Alcohol	40.000	42.584	-6.5	115	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	43.658	-9.1	119	0.00
17	2-Methylphenol	40.000	42.926	-7.3	118	0.00
18	Hexachloroethane	40.000	43.376	-8.4	120	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.763	-6.9	120	0.00
20	3+4-Methylphenols	40.000	42.640	-6.6	118	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	106	0.00
22	Acetophenone	40.000	42.529	-6.3	118	0.00
23 S	Nitrobenzene-d5	80.000	86.189	-7.7	118	0.00
24	Nitrobenzene	40.000	43.626	-9.1	118	0.00
25	Isophorone	40.000	45.910	-14.8	122	0.00
26 C	2-Nitrophenol	40.000	44.097	-10.2	118	0.00
27	2,4-Dimethylphenol	40.000	44.497	-11.2	118	0.00
28	bis(2-Chloroethoxy)methane	40.000	45.087	-12.7	123	0.00
29 C	2,4-Dichlorophenol	40.000	44.810	-12.0	121	0.00
30	1,2,4-Trichlorobenzene	40.000	43.100	-7.8	118	0.00
31	Naphthalene	40.000	43.519	-8.8	119	0.00
32	Benzoic acid	40.000	38.611	3.5	104	0.01
33	4-Chloroaniline	40.000	45.400	-13.5	123	0.00
34 C	Hexachlorobutadiene	40.000	43.489	-8.7	119	0.00
35	Caprolactam	40.000	46.498	-16.2	121	0.01
36 C	4-Chloro-3-methylphenol	40.000	44.329	-10.8	120	0.00
37	2-Methylnaphthalene	40.000	44.451	-11.1	121	0.00
38	1-Methylnaphthalene	40.000	42.749	-6.9	118	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	108	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.252	-3.1	116	0.00
41 P	Hexachlorocyclopentadiene	40.000	33.709	15.7	88	0.00
42 S	2,4,6-Tribromophenol	80.000	83.607	-4.5	118	0.00
43 C	2,4,6-Trichlorophenol	40.000	44.157	-10.4	121	0.00
44	2,4,5-Trichlorophenol	40.000	43.928	-9.8	115	0.00
45 S	2-Fluorobiphenyl	80.000	81.824	-2.3	118	0.00
46	1,1'-Biphenyl	40.000	41.578	-3.9	119	0.00
47	2-Chloronaphthalene	40.000	41.736	-4.3	118	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032522\
 Data File : BF127518.D
 Acq On : 25 Mar 2022 09:48
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 28 05:16:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 22 02:16:33 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.514	-8.8	121	0.00
49	Acenaphthylene	40.000	41.391	-3.5	115	0.00
50	Dimethylphthalate	40.000	43.263	-8.2	122	0.00
51	2,6-Dinitrotoluene	40.000	43.427	-8.6	121	0.00
52 C	Acenaphthene	40.000	42.178	-5.4	117	0.00
53	3-Nitroaniline	40.000	42.899	-7.2	119	0.00
54 P	2,4-Dinitrophenol	40.000	39.332	1.7	109	0.00
55	Dibenzofuran	40.000	44.938	-12.3	118	0.00
56 P	4-Nitrophenol	40.000	37.855	5.4	97	0.01
57	2,4-Dinitrotoluene	40.000	42.014	-5.0	117	0.00
58	Fluorene	40.000	44.398	-11.0	120	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.850	-7.1	122	0.00
60	Diethylphthalate	40.000	43.150	-7.9	124	0.00
61	4-Chlorophenyl-phenylether	40.000	40.487	-1.2	121	0.00
62	4-Nitroaniline	40.000	42.745	-6.9	121	0.00
63	Azobenzene	40.000	42.823	-7.1	122	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.444	-16.1	116	0.00
66 c	n-Nitrosodiphenylamine	40.000	43.906	-9.8	119	0.00
67	4-Bromophenyl-phenylether	40.000	44.543	-11.4	119	0.00
68	Hexachlorobenzene	40.000	44.108	-10.3	119	0.00
69	Atrazine	40.000	44.386	-11.0	115	0.00
70 C	Pentachlorophenol	40.000	43.050	-7.6	114	0.00
71	Phenanthrene	40.000	42.983	-7.5	117	0.00
72	Anthracene	40.000	43.154	-7.9	117	0.00
73	Carbazole	40.000	43.076	-7.7	116	0.00
74	Di-n-butylphthalate	40.000	47.324	-18.3	126	0.00
75 C	Fluoranthene	40.000	42.022	-5.1	113	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	96	0.00
77	Benzydine	40.000	41.503	-3.8	111	0.00
78	Pyrene	40.000	44.836	-12.1	113	0.00
79 S	Terphenyl-d14	80.000	90.864	-13.6	115	0.00
80	Butylbenzylphthalate	40.000	49.962	-24.9	120	0.00
81	Benzo(a)anthracene	40.000	43.343	-8.4	107	0.00
82	3,3'-Dichlorobenzidine	40.000	42.480	-6.2	106	0.00
83	Chrysene	40.000	43.459	-8.6	108	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	51.576	-28.9#	123	0.00
85 c	Di-n-octyl phthalate	40.000	54.945	-37.4#	132	0.00
86 I	Perylene-d12	20.000	20.000	0.0	108	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.808	-7.0	109	0.01
88	Benzo(b)fluoranthene	40.000	41.869	-4.7	112	0.00
89	Benzo(k)fluoranthene	40.000	42.472	-6.2	125	0.00
90 C	Benzo(a)pyrene	40.000	43.885	-9.7	121	0.00
91	Dibenzo(a,h)anthracene	40.000	43.503	-8.8	111	0.00
92	Benzo(g,h,i)perylene	40.000	41.646	-4.1	106	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032522\
Data File : BF127518.D
Acq On : 25 Mar 2022 09:48
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 28 05:16:48 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 22 02:16:33 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1