

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF032619\
 Data File : BF113290.D
 Acq On : 26 Mar 2019 14:25
 Operator : JU/SJ
 Sample : K2081-14
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 S3A(22-28)-S3B(27)COMP

Quant Time: Mar 27 06:52:08 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 03:32:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.70	152	135543	20.00	ng	0.00
21) Naphthalene-d8	7.98	136	526027	20.00	ng	0.00
39) Acenaphthene-d10	9.73	164	255823	20.00	ng	0.00
64) Phenanthrene-d10	11.21	188	512944	20.00	ng	0.00
76) Chrysene-d12	13.83	240	432658	20.00	ng	-0.01
87) Perylene-d12	15.24	264	398923	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.32	112	343880	41.48	ng	0.00
7) Phenol-d6	6.35	99	606362	57.82	ng	-0.01
23) Nitrobenzene-d5	7.26	82	356024	40.17	ng	-0.01
42) 2,4,6-Tribromophenol	10.52	330	277198	87.72	ng	0.00
45) 2-Fluorobiphenyl	9.06	172	854347	43.79	ng	0.00
79) Terphenyl-d14	12.79	244	1392594	70.95	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.45	163	45454	2.398	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF032619\
Data File : BF113290.D
Acq On : 26 Mar 2019 14:25
Operator : JU/SJ
Sample : K2081-14
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
S3A(22-28)-S3B(27)COMP

Quant Time: Mar 27 06:52:08 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 03:32:08 2019
Response via : Initial Calibration

