

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF032619\
 Data File : BF113301.D
 Acq On : 26 Mar 2019 19:57
 Operator : JU/SJ
 Sample : K2081-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 S3A(2-10)COMP

Quant Time: Mar 27 07:19:10 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 03:32:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.70	152	162462	20.00	ng	0.00
21) Naphthalene-d8	7.98	136	608039	20.00	ng	0.00
39) Acenaphthene-d10	9.73	164	275592	20.00	ng	0.00
64) Phenanthrene-d10	11.21	188	512267	20.00	ng	0.00
76) Chrysene-d12	13.83	240	381167	20.00	ng	-0.01
87) Perylene-d12	15.24	264	408160	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.32	112	1097019	110.41	ng	0.00
7) Phenol-d6	6.35	99	1501398	119.45	ng	0.00
23) Nitrobenzene-d5	7.26	82	941642	91.92	ng	0.00
42) 2,4,6-Tribromophenol	10.52	330	263948	77.54	ng	-0.01
45) 2-Fluorobiphenyl	9.06	172	1609367	93.24	ng	0.00
79) Terphenyl-d14	12.79	244	1500988	86.80	ng	0.00
Target Compounds						
10) Phenol	6.36	94	50388	3.511	ng	Qvalue # 70
50) Dimethylphthalate	9.45	163	144296	7.065	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF032619\
Data File : BF113301.D
Acq On : 26 Mar 2019 19:57
Operator : JU/SJ
Sample : K2081-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
S3A(2-10)COMP

Quant Time: Mar 27 07:19:10 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 03:32:08 2019
Response via : Initial Calibration

