

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF032619\
 Data File : BF113305.D
 Acq On : 26 Mar 2019 21:51
 Operator : JU/SJ
 Sample : K2099-07
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SHORE-RD-CLVRT-STLD-2H-A

Manual Integrations
 APPROVED

Sohil
 3/27/2019 3:17:48 PM

Quant Time: Mar 27 07:41:12 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 03:32:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.70	152	153498	20.00	ng	0.00
21) Naphthalene-d8	7.98	136	549498	20.00	ng	0.00
39) Acenaphthene-d10	9.73	164	256460	20.00	ng	0.00
64) Phenanthrene-d10	11.21	188	522182	20.00	ng	0.00
76) Chrysene-d12	13.83	240	354880	20.00	ng	-0.01
87) Perylene-d12	15.23	264	409931	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.32	112	417606	44.49	ng	0.00
7) Phenol-d6	6.35	99	350926	29.55	ng	-0.01
23) Nitrobenzene-d5	7.27	82	908370	98.12	ng	0.00
42) 2,4,6-Tribromophenol	10.52	330	254054	80.20	ng	-0.01
45) 2-Fluorobiphenyl	9.06	172	1550782	97.34	ng	0.00
79) Terphenyl-d14	12.79	244	1412957	87.76	ng	0.00
Target Compounds						
22) Acetophenone	7.10	105	106940	8.534	ng	# 54
31) Naphthalene	8.00	128	97166	3.618	ng	# 96
37) 2-Methylnaphthalene	8.70	142	114950	6.861	ng	97
38) 1-Methylnaphthalene	8.79	142	100063m	6.209	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF032619\
Data File : BF113305.D
Acq On : 26 Mar 2019 21:51
Operator : JU/SJ
Sample : K2099-07
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SHORE-RD-CLVRT-STLD-2H-A

Manual Integrations
APPROVED

Sohil
3/27/2019 3:17:48 PM

Quant Time: Mar 27 07:41:12 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 03:32:08 2019
Response via : Initial Calibration

