

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF032619\
 Data File : BF113311.D
 Acq On : 27 Mar 2019 00:49
 Operator : JU/SJ
 Sample : K2103-05
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MH-C05-SETTLED-3H-A

Quant Time: Mar 27 08:20:34 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 03:32:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.70	152	173092	20.00	ng	0.00
21) Naphthalene-d8	7.98	136	617155	20.00	ng	0.00
39) Acenaphthene-d10	9.73	164	284603	20.00	ng	0.00
64) Phenanthrene-d10	11.21	188	538313	20.00	ng	0.00
76) Chrysene-d12	13.83	240	312800	20.00	ng	-0.01
87) Perylene-d12	15.23	264	377058	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.32	112	366678	34.64	ng	0.00
7) Phenol-d6	6.35	99	297781	22.24	ng	-0.01
23) Nitrobenzene-d5	7.27	82	951114	91.47	ng	0.00
42) 2,4,6-Tribromophenol	10.52	330	249357	70.93	ng	-0.01
45) 2-Fluorobiphenyl	9.06	172	1575065	87.20	ng	0.00
79) Terphenyl-d14	12.79	244	1222011	86.11	ng	0.00
Target Compounds						
25) Isophorone	7.52	82	104001	5.161	ng	Qvalue # 96
30) 1,2,4-Trichlorobenzene	7.93	180	32814	3.402	ng	96
50) Dimethylphthalate	9.45	163	42699	2.024	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF032619\
Data File : BF113311.D
Acq On : 27 Mar 2019 00:49
Operator : JU/SJ
Sample : K2103-05
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MH-C05-SETTLED-3H-A

Quant Time: Mar 27 08:20:34 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 03:32:08 2019
Response via : Initial Calibration

