

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF032619\
 Data File : BF113312.D
 Acq On : 27 Mar 2019 1:18
 Operator : JU/SJ
 Sample : K2099-02
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SHORE-RD-CLVRT-STLD-1H-B

Manual Integrations
 APPROVED

Sohil
 3/27/2019 3:17:50 PM

Quant Time: Mar 27 08:23:51 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 03:32:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.70	152	150396	20.00	ng	0.00
21) Naphthalene-d8	7.99	136	513324	20.00	ng	0.00
39) Acenaphthene-d10	9.73	164	244930	20.00	ng	0.00
64) Phenanthrene-d10	11.21	188	453727	20.00	ng	0.00
76) Chrysene-d12	13.83	240	286652	20.00	ng	-0.01
87) Perylene-d12	15.23	264	337741	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.32	112	397298	43.19	ng	0.00
7) Phenol-d6	6.35	99	320136	27.51	ng	-0.01
23) Nitrobenzene-d5	7.26	82	816072	94.36	ng	0.00
42) 2,4,6-Tribromophenol	10.52	330	246906	81.61	ng	-0.01
45) 2-Fluorobiphenyl	9.06	172	1430458	93.25	ng	0.00
79) Terphenyl-d14	12.79	244	1128640	86.79	ng	0.00
Target Compounds						
10) Phenol	6.36	94	34485	2.596	ng	# 76
31) Naphthalene	8.00	128	52293	2.085	ng	# 88
37) 2-Methylnaphthalene	8.70	142	62813	4.013	ng	97
38) 1-Methylnaphthalene	8.79	142	67648m	4.493	ng	
50) Dimethylphthalate	9.45	163	84737	4.668	ng	100
71) Phenanthrene	11.23	178	48508	2.153	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF032619\
Data File : BF113312.D
Acq On : 27 Mar 2019 1:18
Operator : JU/SJ
Sample : K2099-02
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SHORE-RD-CLVRT-STLD-1H-B

Manual Integrations
APPROVED

Sohil
3/27/2019 3:17:50 PM

Quant Time: Mar 27 08:23:51 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 03:32:08 2019
Response via : Initial Calibration

