

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF032619\
 Data File : BF113298.D
 Acq On : 26 Mar 2019 18:23
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 27 05:00:32 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 03:32:08 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.621	0.608	2.1	105	0.00
3	Pyridine	1.530	1.565	-2.3	117	0.00
4	n-Nitrosodimethylamine	0.680	0.630	7.4	110	0.00
5 S	2-Fluorophenol	1.223	1.144	6.5	107	0.00
6	Aniline	1.884	1.855	1.5	103	0.00
7 S	Phenol-d6	1.547	1.469	5.0	102	0.00
8	2-Chlorophenol	1.420	1.394	1.8	105	0.00
9	Benzaldehyde	1.038	1.034	0.4	106	0.00
10 C	Phenol	1.767	1.696	4.0	103	0.00
11	bis(2-Chloroethyl)ether	1.354	1.298	4.1	104	0.00
12	1,3-Dichlorobenzene	1.531	1.467	4.2	103	0.00
13 C	1,4-Dichlorobenzene	1.478	1.487	-0.6	102	0.00
14	1,2-Dichlorobenzene	1.454	1.406	3.3	102	0.00
15	Benzyl Alcohol	1.021	1.079	-5.7	103	0.00
16	2,2'-oxybis(1-Chloropropane	2.034	2.099	-3.2	102	0.00
17	2-Methylphenol	1.113	1.126	-1.2	101	0.00
18	Hexachloroethane	0.531	0.516	2.8	96	0.00
19 P	n-Nitroso-di-n-propylamine	0.972	0.867	10.8	95	0.00
20	3+4-Methylphenols	1.394	1.275	8.5	96	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
22	Acetophenone	0.456	0.444	2.6	97	0.00
23 S	Nitrobenzene-d5	0.337	0.348	-3.3	97	0.00
24	Nitrobenzene	0.352	0.365	-3.7	95	0.00
25	Isophorone	0.653	0.658	-0.8	98	0.00
26 C	2-Nitrophenol	0.186	0.196	-5.4	100	0.00
27	2,4-Dimethylphenol	0.267	0.279	-4.5	101	0.00
28	bis(2-Chloroethoxy)methane	0.411	0.428	-4.1	102	0.00
29 C	2,4-Dichlorophenol	0.290	0.295	-1.7	99	0.00
30	1,2,4-Trichlorobenzene	0.313	0.317	-1.3	99	0.00
31	Naphthalene	0.977	1.007	-3.1	106	0.00
32	Benzoic acid	0.215	0.227	-5.6	105	0.00
33	4-Chloroaniline	0.381	0.411	-7.9	121	0.00
34 C	Hexachlorobutadiene	0.186	0.195	-4.8	116	0.00
35	Caprolactam	0.093	0.105	-12.9	115	0.00
36 C	4-Chloro-3-methylphenol	0.291	0.322	-10.7	126	0.00
37	2-Methylnaphthalene	0.610	0.677	-11.0	125	0.00
38	1-Methylnaphthalene	0.587	0.662	-12.8	127	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	127	0.00
40	1,2,4,5-Tetrachlorobenzene	0.642	0.643	-0.2	127	0.00
41 P	Hexachlorocyclopentadiene	0.272	0.253	7.0	122	0.00
42 S	2,4,6-Tribromophenol	0.247	0.255	-3.2	128	0.00
43 C	2,4,6-Trichlorophenol	0.418	0.423	-1.2	129	0.00
44	2,4,5-Trichlorophenol	0.463	0.469	-1.3	131	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF032619\
 Data File : BF113298.D
 Acq On : 26 Mar 2019 18:23
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 27 05:00:32 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 03:32:08 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.364	1.306	4.3	123	0.00
46	1,1'-Biphenyl	1.670	1.674	-0.2	128	0.00
47	2-Chloronaphthalene	1.286	1.287	-0.1	127	0.00
48	2-Nitroaniline	0.404	0.424	-5.0	134	0.00
49	Acenaphthylene	2.080	2.137	-2.7	127	0.00
50	Dimethylphthalate	1.482	1.511	-2.0	129	0.00
51	2,6-Dinitrotoluene	0.334	0.350	-4.8	131	0.00
52 C	Acenaphthene	1.171	1.201	-2.6	129	0.00
53	3-Nitroaniline	0.377	0.405	-7.4	134	0.00
54 P	2,4-Dinitrophenol	0.166	0.165	0.6	134	0.00
55	Dibenzofuran	1.761	1.795	-1.9	126	0.00
56 P	4-Nitrophenol	0.269	0.285	-5.9	132	0.00
57	2,4-Dinitrotoluene	0.425	0.453	-6.6	132	0.00
58	Fluorene	1.368	1.372	-0.3	125	0.00
59	2,3,4,6-Tetrachlorophenol	0.391	0.405	-3.6	125	0.00
60	Diethylphthalate	1.454	1.440	1.0	125	0.00
61	4-Chlorophenyl-phenylether	0.667	0.669	-0.3	124	0.00
62	4-Nitroaniline	0.388	0.421	-8.5	134	0.00
63	Azobenzene	1.369	1.392	-1.7	127	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	126	0.00
65	4,6-Dinitro-2-methylphenol	0.109	0.110	-0.9	131	0.00
66 c	n-Nitrosodiphenylamine	0.614	0.611	0.5	128	0.00
67	4-Bromophenyl-phenylether	0.218	0.221	-1.4	127	0.00
68	Hexachlorobenzene	0.253	0.259	-2.4	127	0.00
69	Atrazine	0.194	0.196	-1.0	126	0.00
70 C	Pentachlorophenol	0.132	0.137	-3.8	135	0.00
71	Phenanthrene	0.993	1.015	-2.2	127	0.00
72	Anthracene	1.009	1.034	-2.5	127	0.00
73	Carbazole	0.963	1.007	-4.6	130	0.00
74	Di-n-butylphthalate	1.116	1.103	1.2	122	0.00
75 C	Fluoranthene	1.123	1.109	1.2	125	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	128	0.00
77	Benzidine	0.766	0.793	-3.5	131	0.00
78	Pyrene	1.386	1.410	-1.7	127	0.00
79 S	Terphenyl-d14	0.907	0.883	2.6	121	0.00
80	Butylbenzylphthalate	0.611	0.591	3.3	122	0.00
81	Benzo(a)anthracene	1.145	1.156	-1.0	126	0.00
82	3,3'-Dichlorobenzidine	0.459	0.489	-6.5	130	0.00
83	Chrysene	1.163	1.189	-2.2	128	0.00
84	Bis(2-ethylhexyl)phthalate	0.749	0.700	6.5	117	0.00
85 c	Di-n-octyl phthalate	1.271	1.302	-2.4	124	0.00
86	Indeno(1,2,3-cd)pyrene	0.998	1.039	-4.1	141	0.00
87 I	Perylene-d12	1.000	1.000	0.0	140	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF032619\
Data File : BF113298.D
Acq On : 26 Mar 2019 18:23
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 27 05:00:32 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 03:32:08 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.242	1.289	-3.8	144	0.00
89	Benzo(k)fluoranthene	1.091	1.023	6.2	123	0.00
90 C	Benzo(a)pyrene	1.101	1.082	1.7	138	0.00
91	Dibenzo(a,h)anthracene	0.952	0.907	4.7	140	0.00
92	Benzo(g,h,i)perylene	0.960	0.916	4.6	145	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0