


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032620\
 Data File : BF119533.D
 Acq On : 26 Mar 2020 14:54
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF031820.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Mar 27 07:52:09 2020



AutoFind: Scans 1621, 1622, 1623; Background Corrected with Scan 1616

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.7	76370	PASS
68	69	0.00	2	1.6	939	PASS
69	198	0.00	100	36.3	59282	PASS
70	69	0.00	2	0.2	121	PASS
127	198	10	80	49.4	80669	PASS
197	198	0.00	2	0.6	1043	PASS
198	198	100	100	100.0	163370	PASS
199	198	5	9	6.5	10544	PASS
275	198	10	60	26.5	43269	PASS
365	198	1	100	3.1	5054	PASS
441	198	0.01	100	14.7	24094	PASS
442	198	50	100	88.9	145269	PASS
443	442	15	24	18.9	27388	PASS