


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032719\
 Data File : BF113322.D
 Acq On : 27 Mar 2019 10:06
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032519.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Mar 26 03:32:08 2019



AutoFind: Scans 1596, 1597, 1598; Background Corrected with Scan 1590

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.7	56512	PASS
68	69	0.00	2	1.7	870	PASS
69	198	0.00	100	32.9	51981	PASS
70	69	0.00	2	0.2	103	PASS
127	198	10	80	44.9	71106	PASS
197	198	0.00	2	0.3	510	PASS
198	198	100	100	100.0	158200	PASS
199	198	5	9	6.7	10660	PASS
275	198	10	60	28.4	44856	PASS
365	198	1	100	3.2	5092	PASS
441	198	0.01	100	13.7	21678	PASS
442	198	50	100	83.9	132666	PASS
443	442	15	24	18.8	24891	PASS