


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032919\
 Data File : BF113406.D
 Acq On : 29 Mar 2019 10:40
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032519.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Mar 26 03:32:08 2019



AutoFind: Scans 1596, 1597, 1598; Background Corrected with Scan 1591

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.3	42960	PASS
68	69	0.00	2	1.5	604	PASS
69	198	0.00	100	31.5	40675	PASS
70	69	0.00	2	0.6	227	PASS
127	198	10	80	44.5	57408	PASS
197	198	0.00	2	0.4	558	PASS
198	198	100	100	100.0	129061	PASS
199	198	5	9	6.7	8606	PASS
275	198	10	60	28.5	36762	PASS
365	198	1	100	3.3	4214	PASS
441	198	0.01	100	13.8	17839	PASS
442	198	50	100	87.1	112458	PASS
443	442	15	24	19.2	21571	PASS