

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF033020\
 Data File : BF119643.D
 Acq On : 30 Mar 2020 19:53
 Operator : MA/SJ
 Sample : L2042-02
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-02-1.0-1.5

Manual Integrations
 APPROVED

mohammad
 3/31/2020 3:16:40 PM

Quant Time: Mar 31 01:40:13 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF031820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 27 07:52:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.83	152	292775	20.00	ng	0.00
21) Naphthalene-d8	8.12	136	1034425	20.00	ng	0.00
39) Acenaphthene-d10	9.88	164	466958	20.00	ng	0.00
64) Phenanthrene-d10	11.37	188	720164	20.00	ng	0.00
76) Chrysene-d12	14.01	240	518644	20.00	ng	0.00
87) Perylene-d12	15.48	264	457780	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	1729503	94.04	ng	0.00
7) Phenol-d6	6.46	99	2261875	96.28	ng	0.00
23) Nitrobenzene-d5	7.39	82	1076383	56.55	ng	-0.01
42) 2,4,6-Tribromophenol	10.67	330	338578	70.28	ng	0.00
45) 2-Fluorobiphenyl	9.20	172	2401817	76.20	ng	0.00
79) Terphenyl-d14	12.96	244	1960166	74.05	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.59	163	263943	7.958	ng	98
71) Phenanthrene	11.39	178	118706	3.179	ng	99
75) Fluoranthene	12.58	202	206207	5.102	ng	99
78) Pyrene	12.81	202	182531	4.288	ng	98
81) Benzo(a)anthracene	14.00	228	92133	2.732	ng	95
83) Chrysene	14.04	228	81067	2.329	ng	96
88) Benzo(b)fluoranthene	15.05	252	101034m	3.304	ng	
90) Benzo(a)pyrene	15.42	252	66911	2.408	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF033020\
Data File : BF119643.D
Acq On : 30 Mar 2020 19:53
Operator : MA/SJ
Sample : L2042-02
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-02-1.0-1.5

Manual Integrations
APPROVED

mohammad
3/31/2020 3:16:40 PM

Quant Time: Mar 31 01:40:13 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF031820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 27 07:52:09 2020
Response via : Initial Calibration

