

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF040119\
 Data File : BF113494.D
 Acq On : 2 Apr 2019 6:37
 Operator : JU/SJ
 Sample : K2056-03DL 2X
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Apr 02 07:04:15 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 03:32:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.71	152	132528	20.00	ng	0.00
21) Naphthalene-d8	7.99	136	499839	20.00	ng	0.00
39) Acenaphthene-d10	9.74	164	219409	20.00	ng	0.00
64) Phenanthrene-d10	11.22	188	395006	20.00	ng	0.00
76) Chrysene-d12	13.85	240	363840	20.00	ng	0.00
87) Perylene-d12	15.26	264	249175	20.00	ng	0.01

System Monitoring Compounds

5) 2-Fluorophenol	5.32	112	374693	46.23	ng	0.00
7) Phenol-d6	6.35	99	517008	50.42	ng	0.00
23) Nitrobenzene-d5	7.27	82	266983	31.70	ng	0.00
42) 2,4,6-Tribromophenol	10.53	330	116429	42.96	ng	0.00
45) 2-Fluorobiphenyl	9.06	172	581831	30.19	ng	0.00
79) Terphenyl-d14	12.80	244	574120	34.78	ng	0.00

Target Compounds


						Qvalue
19) n-Nitroso-di-n-propylamine	7.27	70	34767	5.398	ng	# 77
25) Isophorone	7.27	82	266980	16.359	ng	# 64
49) Acenaphthylene	9.60	152	49721	2.179	ng	99
51) 2,6-Dinitrotoluene	9.74	165	28074	7.653	ng	# 26
52) Acenaphthene	9.77	154	34868	2.714	ng	99
56) 4-Nitrophenol	9.95	139	7486	2.538	ng	# 7
57) 2,4-Dinitrotoluene	9.74	165	28074	6.018	ng	# 23
58) Fluorene	10.29	166	31473	2.098	ng	97
71) Phenanthrene	11.24	178	340685	17.368	ng	98
72) Anthracene	11.29	178	128865	6.468	ng	98
73) Carbazole	11.45	167	82389	4.333	ng	98
75) Fluoranthene	12.43	202	1594636	71.909	ng	100
78) Pyrene	12.66	202	1362697	54.048	ng	99
81) Benzo(a)anthracene	13.84	228	386469	18.556	ng	97
83) Chrysene	13.87	228	819901	38.758	ng	99
86) Indeno(1,2,3-cd)pyrene	16.62	276	78592	4.329	ng	95
88) Benzo(b)fluoranthene	14.86	252	724395	46.800	ng	# 98
89) Benzo(k)fluoranthene	14.86	252	724395	53.291	ng	# 97
90) Benzo(a)pyrene	15.20	252	120275	8.765	ng	97
92) Benzo(a,h,i)perylene	17.03	276	62067	5.188	ng	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF040119\
Data File : BF113494.D
Acq On : 2 Apr 2019 6:37
Operator : JU/SJ
Sample : K2056-03DL 2X
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: Apr 02 07:04:15 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 03:32:08 2019
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.71 min Scan# 786
Delta R.T. 0.01 min
Lab File: BF113494.D
Acq: 2 Apr 2019 6:37

Instrument :
BNA_F
ClientSampleId :

Tot Ion:152 Resp: 132528
Ion Ratio Lower Upper
152 100
150 151.8 124.7 187.1
115 52.1 46.0 69.0




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 46.229 ng
RT: 5.32 min Scan# 550
Delta R.T. 0.01 min
Lab File: BF113494.D
Acq: 2 Apr 2019 6:37

Tgt Ion:112 Resp: 374693
Ion Ratio Lower Upper
112 100
64 58.9 49.2 73.8
63 28.8 24.9 37.3



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 50.423 ng

RT: 6.35 min Scan# 725

Delta R.T. -0.01 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37

Instrument :

BNA_F

ClientSampleId :





Tot Ion: 99 Resp: 517008

Ion Ratio Lower Upper

99 100

42 17.7 14.8 22.2

71 30.2 24.9 37.3



#19

n-Nitroso-di-n-propylamine

Concen: 5.398 ng

RT: 7.27 min Scan# 881

Delta R.T. 0.14 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37

Tgt Ion: 70 Resp: 34767




Ion Ratio Lower Upper


70 100

42 48.4 48.4 72.6

101 0.0 7.8 11.8#

130 2.5 18.2 27.4#





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 7.99 min Scan# 1003

Delta R.T. -0.00 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37

Instrument :

BNA_F

ClientSampleId :

Tot Ion:136 Resp: 499839


Ion Ratio Lower Upper

136 100

137 11.3 8.9 13.3

54 8.7 7.0 10.6

68 6.7 4.8 7.2






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1



#23

Nitrobenzene-d5

Concen: 31.703 ng

RT: 7.27 min Scan# 881

Delta R.T. 0.00 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37


Tgt Ion: 82 Resp: 266983

Ion Ratio Lower Upper

82 100

128 46.7 36.5 54.7



54 50.5 42.2 63.2




Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BF1





#25

Isophorone

Concen: 16.359 ng

RT: 7.27 min Scan# 881

Delta R.T. -0.26 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 82 Resp: 266980

Ion Ratio Lower Upper

82 100

95 0.0 5.5 8.3#




138 0.0 15.2 22.8#



Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 95.00 (94.70 to 95.70): BF1

Ion 138.00 (137.70 to 138.70): BF1



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.74 min Scan# 1301

Delta R.T. -0.00 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37


Tgt Ion:164 Resp: 219409

Ion Ratio Lower Upper

164 100

162 101.4 81.8 122.6



160 43.9 36.4 54.6




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E





#42

2,4,6-Tribromophenol

Concen: 42.960 ng

RT: 10.53 min Scan# 1435

Delta R.T. 0.00 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37

Instrument :

BNA_F

ClientSampleId :


Tot Ion:330 Resp: 116429

Ion Ratio Lower Upper




330 100

332 95.8 77.8 116.6

141 28.1 22.7 34.1



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E



#45

2-Fluorobiphenyl

Concen: 30.191 ng

RT: 9.06 min Scan# 1186

Delta R.T. 0.00 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37


Tgt Ion:172 Resp: 581831

Ion Ratio Lower Upper



172 100


171 35.2 29.0 43.6

170 24.2 19.5 29.3



Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E





#49

Acenaphthylene

Concen: 2.179 ng

RT: 9.60 min Scan# 1277

Delta R.T. -0.00 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37

Instrument :

BNA_F

ClientSampleId :


Tot Ion:152 Resp: 49721

Ion Ratio Lower Upper

152 100

151 20.1 16.2 24.4




153 13.4 11.0 16.6



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#51

2,6-Dinitrotoluene

Concen: 7.653 ng

RT: 9.74 min Scan# 1301

Delta R.T. 0.21 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37


Tgt Ion:165 Resp: 28074

Ion Ratio Lower Upper

165 100

63 1.4 43.4 65.0#



89 1.3 43.1 64.7#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 89.00 (88.70 to 89.70): BF1





#52

Acenaphthene

Concen: 2.714 ng

RT: 9.77 min Scan# 1306

Delta R.T. -0.01 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37

Instrument :

BNA_F

ClientSampled :


Tot Ion:154 Resp: 34868

Ion Ratio Lower Upper

154 100

153 111.6 89.5 134.3




152 55.1 43.3 64.9



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#56

4-Nitrophenol

Concen: 2.538 ng

RT: 9.95 min Scan# 1336

Delta R.T. 0.08 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37


Tgt Ion:139 Resp: 7486

Ion Ratio Lower Upper

139 100

109 0.0 46.9 86.9#



65 0.0 77.2 117.2#




Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF1









#57
 2,4-Dinitrotoluene
 Concen: 6.018 ng
 RT: 9.74 min Scan# 1301
 Delta R.T. -0.19 min
 Lab File: BF113494.D
 Acq: 2 Apr 2019 6:37




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:	165	Resp:	28074
Ion Ratio	Lower	Upper	
165	100		
63	1.4	37.5	56.3#
89	1.3	57.4	86.2#
182	0.0	2.0	3.0#



#58
 Fluorene
 Concen: 2.098 ng
 RT: 10.29 min Scan# 1394
 Delta R.T. -0.00 min
 Lab File: BF113494.D
 Acq: 2 Apr 2019 6:37

Tgt Ion:	166	Resp:	31473
Ion Ratio	Lower	Upper	
166	100		
165	97.9	76.2	114.2
167	15.8	10.7	16.1





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.22 min Scan# 1552

Delta R.T. -0.00 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37


Instrument :

BNA_F

ClientSampleId :

Tot Ion:188 Resp: 395006


Ion	Ratio	Lower	Upper
188	100		
94	9.6	7.0	10.6
80	10.1	7.7	11.5




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1



Time-->



#71

Phenanthrene

Concen: 17.368 ng

RT: 11.24 min Scan# 1556

Delta R.T. -0.00 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37



Tgt Ion:178 Resp: 340685

Ion	Ratio	Lower	Upper
178	100		
176	18.6	16.2	24.2
179	15.5	12.7	19.1


Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E







Time-->



#72
 Anthracene
 Concen: 6.468 ng
 RT: 11.29 min Scan# 1565
 Delta R.T. -0.00 min
 Lab File: BF113494.D
 Acq: 2 Apr 2019 6:37




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:178 Resp: 128865
 Ion Ratio Lower Upper
 178 100
 176 18.3 15.4 23.2
 179 16.5 12.8 19.2



#73
 Carbazole
 Concen: 4.333 ng
 RT: 11.45 min Scan# 1592
 Delta R.T. 0.01 min
 Lab File: BF113494.D
 Acq: 2 Apr 2019 6:37

Tgt Ion:167 Resp: 82389
 Ion Ratio Lower Upper
 167 100
 166 20.5 17.4 26.0
 139 13.1 10.4 15.6





#75

Fluoranthene

Concen: 71.909 ng

RT: 12.43 min Scan# 1759

Delta R.T. 0.01 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37

Instrument :

BNA_F

ClientSampleId :


Tot Ion:202 Resp: 1594636

Ion Ratio Lower Upper

202 100

101 11.5 0.0 31.4




203 18.0 0.0 38.0



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.85 min Scan# 2000

Delta R.T. 0.01 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37


Tgt Ion:240 Resp: 363840

Ion Ratio Lower Upper

240 100

120 10.2 8.8 13.2



236 25.7 20.9 31.3




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E









#78
Pvrene
Concen: 54.048 ng
RT: 12.66 min Scan# 1797
Delta R.T. 0.01 min
Lab File: BF113494.D
Acq: 2 Apr 2019 6:37




Instrument :
BNA_F
ClientSampleId :


Tot Ion:202 Resp: 1362697
Ion Ratio Lower Upper
202 100
200 21.2 16.8 25.2
203 18.3 14.4 21.6



#79
Terphenyl-d14
Concen: 34.781 ng
RT: 12.80 min Scan# 1821
Delta R.T. 0.00 min
Lab File: BF113494.D
Acq: 2 Apr 2019 6:37

Tgt Ion:244 Resp: 574120
Ion Ratio Lower Upper
244 100
212 5.9 5.5 8.3
122 9.0 8.6 13.0





#81

Benzo(a)anthracene

Concen: 18.556 ng

RT: 13.84 min Scan# 1998

Delta R.T. 0.01 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37

Instrument :

BNA_F

ClientSampleId :


Tot Ion:228 Resp: 386469

Ion Ratio Lower Upper

228 100

226 29.7 21.8 32.8




229 21.1 16.2 24.4



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#83

Chrysene

Concen: 38.758 ng

RT: 13.87 min Scan# 2004

Delta R.T. -0.00 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37


Tgt Ion:228 Resp: 819901

Ion Ratio Lower Upper

228 100

226 30.0 24.3 36.5



229 21.5 16.3 24.5




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#86
 Indeno(1,2,3-cd)pyrene
 Concen: 4.329 ng
 RT: 16.62 min Scan# 2471
 Delta R.T. 0.02 min
 Lab File: BF113494.D
 Acq: 2 Apr 2019 6:37




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:276 Resp: 78592
 Ion Ratio Lower Upper
 276 100
 138 19.8 18.1 27.1
 277 22.9 20.2 30.4



#87
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.26 min Scan# 2239
 Delta R.T. 0.01 min
 Lab File: BF113494.D
 Acq: 2 Apr 2019 6:37

Tgt Ion:264 Resp: 249175
 Ion Ratio Lower Upper
 264 100
 260 23.3 19.1 28.7
 265 21.8 17.1 25.7





#88

Benzo(b)fluoranthene

Concen: 46.800 ng

RT: 14.86 min Scan# 2171

Delta R.T. 0.01 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37

Instrument :

BNA_F

ClientSampleId :


Tot Ion:252 Resp: 724395

Ion Ratio Lower Upper

252 100

253 22.6 18.0 27.0




125 11.4 7.6 11.4#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#89

Benzo(k)fluoranthene

Concen: 53.291 ng

RT: 14.86 min Scan# 2171

Delta R.T. -0.02 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37


Tgt Ion:252 Resp: 724395

Ion Ratio Lower Upper

252 100

253 22.6 17.7 26.5



125 11.4 7.3 10.9#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#90

Benzo(a)pyrene

Concen: 8.765 ng

RT: 15.20 min Scan# 2229

Delta R.T. 0.01 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37

Instrument :

BNA_F

ClientSampled :


Tot Ion:252 Resp: 120275

Ion Ratio Lower Upper

252 100

253 24.3 17.9 26.9




125 11.6 8.8 13.2



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#92

Benzo(g,h,i)perylene

Concen: 5.188 ng

RT: 17.03 min Scan# 2541

Delta R.T. 0.03 min

Lab File: BF113494.D

Acq: 2 Apr 2019 6:37


Tgt Ion:276 Resp: 62067

Ion Ratio Lower Upper

276 100

277 26.4 18.8 28.2

138 22.9 15.6 23.4



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

