

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF040119\
 Data File : BF113473.D
 Acq On : 1 Apr 2019 18:22
 Operator : JU/SJ
 Sample : K2185-02
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 P-4-OW(0-10)

Manual Integrations
 APPROVED

Sohil
 4/2/2019 3:49:30 PM

Quant Time: Apr 02 02:08:52 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 03:32:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.70	152	143559	20.00	ng	0.00
21) Naphthalene-d8	7.99	136	535904	20.00	ng	0.00
39) Acenaphthene-d10	9.73	164	242414	20.00	ng	0.00
64) Phenanthrene-d10	11.22	188	430444	20.00	ng	0.00
76) Chrysene-d12	13.84	240	383611	20.00	ng	0.00
87) Perylene-d12	15.26	264	325186	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.32	112	840015	95.68	ng	0.00
7) Phenol-d6	6.36	99	1086310	97.81	ng	0.00
23) Nitrobenzene-d5	7.27	82	669260	74.12	ng	0.00
42) 2,4,6-Tribromophenol	10.53	330	275117	91.88	ng	0.00
45) 2-Fluorobiphenyl	9.06	172	1008545	60.03	ng	0.00
79) Terphenyl-d14	12.79	244	902259	51.84	ng	0.00
Target Compounds						Qvalue
50) Dimethylphthalate	9.46	163	74856	4.167	ng	99
71) Phenanthrene	11.24	178	270503	12.655	ng	99
72) Anthracene	11.29	178	76205	3.510	ng	98
74) Di-n-butylphthalate	11.77	149	62528	2.603	ng	99
75) Fluoranthene	12.42	202	558733	23.121	ng	99
78) Pyrene	12.65	202	534962	20.124	ng	99
81) Benzo(a)anthracene	13.83	228	330903	15.069	ng	99
83) Chrysene	13.87	228	314859	14.117	ng	98
86) Indeno(1,2,3-cd)pyrene	16.62	276	131710	6.881	ng	97
88) Benzo(b)fluoranthene	14.86	252	415656m	20.577	ng	
89) Benzo(k)fluoranthene	14.88	252	126315m	7.120	ng	
90) Benzo(a)pyrene	15.20	252	263207	14.697	ng	99
91) Dibenzo(a,h)anthracene	16.62	278	39168	2.529	ng	# 77
92) Benzo(g,h,i)perylene	17.03	276	129291	8.281	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF040119\
Data File : BF113473.D
Acq On : 1 Apr 2019 18:22
Operator : JU/SJ
Sample : K2185-02
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
P-4-OW(0-10)

Manual Integrations
APPROVED

Sohil
4/2/2019 3:49:30 PM

Quant Time: Apr 02 02:08:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 03:32:08 2019
Response via : Initial Calibration

