

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF040319\
 Data File : BF113550.D
 Acq On : 3 Apr 2019 18:09
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 04 09:21:46 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF040319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 03 18:55:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.555	0.569	-2.5	96	0.00
3	Pyridine	1.454	1.481	-1.9	93	0.00
4	n-Nitrosodimethylamine	0.618	0.628	-1.6	93	0.00
5 S	2-Fluorophenol	1.173	1.197	-2.0	94	0.00
6	Aniline	1.752	1.849	-5.5	94	0.00
7 S	Phenol-d6	1.446	1.484	-2.6	94	0.00
8	2-Chlorophenol	1.357	1.388	-2.3	94	0.00
9	Benzaldehyde	0.860	0.921	-7.1	93	0.00
10 C	Phenol	1.651	1.694	-2.6	94	0.00
11	bis(2-Chloroethyl)ether	1.284	1.300	-1.2	93	0.00
12	1,3-Dichlorobenzene	1.461	1.502	-2.8	95	0.00
13 C	1,4-Dichlorobenzene	1.477	1.506	-2.0	94	0.00
14	1,2-Dichlorobenzene	1.333	1.389	-4.2	95	0.00
15	Benzyl Alcohol	0.977	1.028	-5.2	92	0.00
16	2,2'-oxybis(1-Chloropropane	1.999	2.045	-2.3	93	0.00
17	2-Methylphenol	1.052	1.076	-2.3	94	0.00
18	Hexachloroethane	0.538	0.547	-1.7	93	0.00
19 P	n-Nitroso-di-n-propylamine	0.891	0.896	-0.6	92	0.00
20	3+4-Methylphenols	1.274	1.318	-3.5	93	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
22	Acetophenone	0.439	0.440	-0.2	92	0.00
23 S	Nitrobenzene-d5	0.341	0.344	-0.9	93	0.00
24	Nitrobenzene	0.352	0.358	-1.7	93	0.00
25	Isophorone	0.658	0.651	1.1	92	0.00
26 C	2-Nitrophenol	0.189	0.194	-2.6	94	0.00
27	2,4-Dimethylphenol	0.271	0.269	0.7	92	0.00
28	bis(2-Chloroethoxy)methane	0.416	0.422	-1.4	93	0.00
29 C	2,4-Dichlorophenol	0.289	0.295	-2.1	93	0.00
30	1,2,4-Trichlorobenzene	0.314	0.317	-1.0	92	0.00
31	Naphthalene	0.969	0.986	-1.8	93	0.00
32	Benzoic acid	0.209	0.201	3.8	87	0.00
33	4-Chloroaniline	0.384	0.399	-3.9	92	0.00
34 C	Hexachlorobutadiene	0.192	0.195	-1.6	93	0.00
35	Caprolactam	0.092	0.092	0.0	89	0.00
36 C	4-Chloro-3-methylphenol	0.291	0.295	-1.4	92	0.00
37	2-Methylnaphthalene	0.637	0.644	-1.1	92	0.00
38	1-Methylnaphthalene	0.615	0.629	-2.3	93	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
40	1,2,4,5-Tetrachlorobenzene	0.647	0.657	-1.5	92	0.00
41 P	Hexachlorocyclopentadiene	0.192	0.171	10.9	90	0.00
42 S	2,4,6-Tribromophenol	0.240	0.240	0.0	87	0.00
43 C	2,4,6-Trichlorophenol	0.408	0.409	-0.2	91	0.00
44	2,4,5-Trichlorophenol	0.438	0.444	-1.4	92	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF040319\
 Data File : BF113550.D
 Acq On : 3 Apr 2019 18:09
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 04 09:21:46 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF040319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 03 18:55:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.242	1.319	-6.2	92	0.00
46	1,1'-Biphenyl	1.650	1.671	-1.3	92	0.00
47	2-Chloronaphthalene	1.250	1.268	-1.4	92	0.00
48	2-Nitroaniline	0.382	0.390	-2.1	92	0.00
49	Acenaphthylene	2.032	2.084	-2.6	92	0.00
50	Dimethylphthalate	1.474	1.448	1.8	89	0.00
51	2,6-Dinitrotoluene	0.325	0.330	-1.5	91	0.00
52 C	Acenaphthene	1.163	1.176	-1.1	92	0.00
53	3-Nitroaniline	0.369	0.370	-0.3	90	0.00
54 P	2,4-Dinitrophenol	0.152	0.137	9.9	77	0.00
55	Dibenzofuran	1.706	1.737	-1.8	91	0.00
56 P	4-Nitrophenol	0.229	0.219	4.4	85	0.00
57	2,4-Dinitrotoluene	0.393	0.399	-1.5	90	0.00
58	Fluorene	1.349	1.375	-1.9	92	0.00
59	2,3,4,6-Tetrachlorophenol	0.379	0.386	-1.8	91	0.00
60	Diethylphthalate	1.421	1.414	0.5	89	0.00
61	4-Chlorophenyl-phenylether	0.664	0.679	-2.3	91	0.00
62	4-Nitroaniline	0.376	0.377	-0.3	90	0.00
63	Azobenzene	1.346	1.359	-1.0	91	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
65	4,6-Dinitro-2-methylphenol	0.105	0.104	1.0	88	0.00
66 c	n-Nitrosodiphenylamine	0.614	0.621	-1.1	90	0.00
67	4-Bromophenyl-phenylether	0.224	0.226	-0.9	89	0.00
68	Hexachlorobenzene	0.257	0.261	-1.6	90	0.00
69	Atrazine	0.194	0.196	-1.0	89	0.00
70 C	Pentachlorophenol	0.121	0.115	5.0	85	0.00
71	Phenanthrene	0.994	1.007	-1.3	89	0.00
72	Anthracene	1.011	1.035	-2.4	90	0.00
73	Carbazole	0.946	0.967	-2.2	90	0.00
74	Di-n-butylphthalate	1.125	1.135	-0.9	87	0.00
75 C	Fluoranthene	1.065	1.089	-2.3	89	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	85	0.00
77	Benzidine	0.744	0.875	-17.6	101	0.00
78	Pyrene	1.522	1.579	-3.7	88	0.00
79 S	Terphenyl-d14	0.982	1.041	-6.0	90	0.00
80	Butylbenzylphthalate	0.619	0.621	-0.3	84	0.00
81	Benzo(a)anthracene	1.154	1.171	-1.5	85	0.00
82	3,3'-Dichlorobenzidine	0.444	0.452	-1.8	83	0.00
83	Chrysene	1.170	1.177	-0.6	85	0.00
84	Bis(2-ethylhexyl)phthalate	0.749	0.741	1.1	83	0.00
85 c	Di-n-octyl phthalate	1.217	1.167	4.1	81	0.00
86	Indeno(1,2,3-cd)pyrene	1.080	1.053	2.5	95	0.00
87 I	Perylene-d12	1.000	1.000	0.0	86	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF040319\
Data File : BF113550.D
Acq On : 3 Apr 2019 18:09
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 04 09:21:46 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF040319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 03 18:55:06 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.224	1.232	-0.7	88	0.00
89	Benzo(k)fluoranthene	1.099	1.143	-4.0	84	0.00
90 C	Benzo(a)pyrene	1.088	1.109	-1.9	88	0.00
91	Dibenzo(a,h)anthracene	1.017	1.064	-4.6	95	0.00
92	Benzo(g,h,i)perylene	1.038	1.081	-4.1	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0