

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF040320\
 Data File : BF119734.D
 Acq On : 3 Apr 2020 22:59
 Operator : MA/SJ
 Sample : L2133-04
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SMALL-HOUSE-WALLS

Quant Time: Apr 04 00:50:27 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF031820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 11:59:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	204109	20.00	ng	-0.01
21) Naphthalene-d8	8.09	136	759843	20.00	ng	-0.02
39) Acenaphthene-d10	9.85	164	370760	20.00	ng	-0.02
64) Phenanthrene-d10	11.33	188	649029	20.00	ng	-0.02
76) Chrysene-d12	13.97	240	385419	20.00	ng	-0.03
87) Perylene-d12	15.42	264	403515	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	783961	61.14	ng	0.00
7) Phenol-d6	6.43	99	979736	59.82	ng	-0.01
23) Nitrobenzene-d5	7.36	82	588072	42.06	ng	-0.02
42) 2,4,6-Tribromophenol	10.63	330	237691	62.14	ng	-0.02
45) 2-Fluorobiphenyl	9.16	172	1155454	46.17	ng	-0.02
79) Terphenyl-d14	12.92	244	1051817	53.47	ng	-0.02
Target Compounds						
50) Dimethylphthalate	9.56	163	59841	2.272	ng	Qvalue 100
84) Bis(2-ethylhexyl)phthalate	13.96	149	104261	6.836	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF040320\
Data File : BF119734.D
Acq On : 3 Apr 2020 22:59
Operator : MA/SJ
Sample : L2133-04
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SMALL-HOUSE-WALLS

Quant Time: Apr 04 00:50:27 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF031820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 01 11:59:45 2020
Response via : Initial Calibration

