

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF040620\
 Data File : BF119812.D
 Acq On : 7 Apr 2020 10:48
 Operator : MA/SJ
 Sample : L2147-01
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-02-056-A

Manual Integrations
 APPROVED

mohammad
 4/8/2020 8:12:46 AM

Quant Time: Apr 07 11:20:05 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF031820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 06 11:15:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.78	152	172706	20.00	ng	0.00
21) Naphthalene-d8	8.07	136	640769	20.00	ng	0.00
39) Acenaphthene-d10	9.82	164	308334	20.00	ng	-0.01
64) Phenanthrene-d10	11.31	188	554347	20.00	ng	0.00
76) Chrysene-d12	13.95	240	387656	20.00	ng	0.00
87) Perylene-d12	15.40	264	381200	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.39	112	1041148	95.97	ng	0.00
7) Phenol-d6	6.41	99	1328039	95.83	ng	0.00
23) Nitrobenzene-d5	7.35	82	798084	67.69	ng	0.00
42) 2,4,6-Tribromophenol	10.62	330	331551	104.23	ng	0.00
45) 2-Fluorobiphenyl	9.15	172	1382587	66.43	ng	0.00
79) Terphenyl-d14	12.90	244	1359693	68.72	ng	0.00
Target Compounds						
15) Benzyl Alcohol	6.92	79	42628	4.089	ng	96
50) Dimethylphthalate	9.54	163	264715	12.087	ng	98
71) Phenanthrene	11.33	178	223189	7.765	ng	98
75) Fluoranthene	12.52	202	404716	13.010	ng	96
78) Pyrene	12.75	202	355929	11.188	ng	97
81) Benzo(a)anthracene	13.94	228	169903	6.740	ng	98
83) Chrysene	13.97	228	156384	6.011	ng	97
84) Bis(2-ethylhexyl)phthalate	13.94	149	92649	6.040	ng	99
86) Indeno(1,2,3-cd)pyrene	16.82	276	83889	3.424	ng	94
88) Benzo(b)fluoranthene	14.98	252	208229m	8.177	ng	
89) Benzo(k)fluoranthene	15.00	252	60390m	2.491	ng	
90) Benzo(a)pyrene	15.33	252	137138	5.926	ng	99
92) Benzo(a,h,i)perylene	17.24	276	78980	3.884	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF040620\
Data File : BF119812.D
Acq On : 7 Apr 2020 10:48
Operator : MA/SJ
Sample : L2147-01
Misc :
ALS Vial : 50 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FK-GF-02-056-A

Manual Integrations
APPROVED

mohammad
4/8/2020 8:12:46 AM

Quant Time: Apr 07 11:20:05 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF031820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 06 11:15:16 2020
Response via : Initial Calibration

