

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF040820\
 Data File : BF119876.D
 Acq On : 9 Apr 2020 18:18
 Operator : MA/SJ
 Sample : L2164-02
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ROLL-OFF-COMP

Quant Time: Apr 10 02:11:14 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF040820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 09 01:21:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.77	152	171045	20.00	ng	0.00
21) Naphthalene-d8	8.05	136	561842	20.00	ng	0.00
39) Acenaphthene-d10	9.81	164	260536	20.00	ng	0.00
64) Phenanthrene-d10	11.29	188	424630	20.00	ng	0.00
76) Chrysene-d12	13.93	240	311885	20.00	ng	0.00
87) Perylene-d12	15.37	264	285024	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.38	112	900250	90.96	ng	0.00
7) Phenol-d6	6.40	99	1252094	98.18	ng	0.00
23) Nitrobenzene-d5	7.33	82	780140	71.83	ng	0.00
42) 2,4,6-Tribromophenol	10.60	330	250746	78.61	ng	0.00
45) 2-Fluorobiphenyl	9.13	172	1173898	63.97	ng	0.00
79) Terphenyl-d14	12.89	244	931094	50.23	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.53	163	133587	6.629	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF040820\
Data File : BF119876.D
Acq On : 9 Apr 2020 18:18
Operator : MA/SJ
Sample : L2164-02
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ROLL-OFF-COMP

Quant Time: Apr 10 02:11:14 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF040820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 09 01:21:19 2020
Response via : Initial Calibration

