

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF041320\
 Data File : BF119940.D
 Acq On : 14 Apr 2020 7:32
 Operator : MA/SJ
 Sample : L2197-01
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CHASSIE-WASH-PIT

Quant Time: Apr 14 09:07:48 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF040820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 13 13:37:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.73	152	213912	20.00	ng	0.00
21) Naphthalene-d8	8.02	136	847820	20.00	ng	-0.01
39) Acenaphthene-d10	9.77	164	454024	20.00	ng	0.00
64) Phenanthrene-d10	11.26	188	896108	20.00	ng	0.00
76) Chrysene-d12	13.90	240	701509	20.00	ng	-0.01
87) Perylene-d12	15.32	264	607459	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.36	112	1688643	136.42	ng	0.02
7) Phenol-d6	6.37	99	1920927	120.44	ng	0.00
23) Nitrobenzene-d5	7.30	82	1465319	89.41	ng	-0.01
42) 2,4,6-Tribromophenol	10.57	330	724521	130.34	ng	0.00
45) 2-Fluorobiphenyl	9.10	172	2759794	86.30	ng	-0.01
79) Terphenyl-d14	12.86	244	3251803	78.00	ng	0.00
Target Compounds						
20) 3+4-Methylphenols	7.14	107	256823	17.014	ng	Qvalue # 61
50) Dimethylphthalate	9.49	163	254991	7.261	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF041320\
Data File : BF119940.D
Acq On : 14 Apr 2020 7:32
Operator : MA/SJ
Sample : L2197-01
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CHASSIE-WASH-PIT

Quant Time: Apr 14 09:07:48 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF040820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 13 13:37:50 2020
Response via : Initial Calibration

