

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF041520\
 Data File : BF120021.D
 Acq On : 16 Apr 2020 4:40
 Operator : MA/SJ
 Sample : L2261-06
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 NCQ-632-X

Quant Time: Apr 16 05:40:27 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF040820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 15 11:01:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.72	152	160422	20.00	ng	0.00
21) Naphthalene-d8	8.00	136	627589	20.00	ng	-0.01
39) Acenaphthene-d10	9.76	164	341270	20.00	ng	0.00
64) Phenanthrene-d10	11.25	188	653120	20.00	ng	0.00
76) Chrysene-d12	13.89	240	479986	20.00	ng	-0.01
87) Perylene-d12	15.31	264	360374	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.00	112	0	0.00	ng	
7) Phenol-d6	6.36	99	2471	0.21	ng	-0.01
23) Nitrobenzene-d5	7.29	82	1054980	86.96	ng	0.00
42) 2,4,6-Tribromophenol	10.55	330	120	0.03	ng	-0.01
45) 2-Fluorobiphenyl	9.09	172	2075026	86.33	ng	0.00
79) Terphenyl-d14	12.84	244	2255172	79.06	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.48	163	172292	6.527	ng	99
74) Di-n-butylphthalate	11.82	149	130717	2.961	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF041520\
Data File : BF120021.D
Acq On : 16 Apr 2020 4:40
Operator : MA/SJ
Sample : L2261-06
Misc :
ALS Vial : 59 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
NCQ-632-X

Quant Time: Apr 16 05:40:27 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF040820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 15 11:01:23 2020
Response via : Initial Calibration

