

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF041619\
 Data File : BF113794.D
 Acq On : 16 Apr 2019 15:27
 Operator : JU/SJ
 Sample : K2422-03
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP1-S

Quant Time: Apr 16 18:08:15 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF040319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 16 17:22:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.67	152	106929	20.00	ng	0.00
21) Naphthalene-d8	7.96	136	384896	20.00	ng	0.00
39) Acenaphthene-d10	9.70	164	164623	20.00	ng	0.00
64) Phenanthrene-d10	11.19	188	280809	20.00	ng	0.00
76) Chrysene-d12	13.83	240	270123	20.00	ng	0.00
87) Perylene-d12	15.24	264	326959	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.29	112	656568	104.70	ng	0.00
7) Phenol-d6	6.33	99	854299	110.54	ng	0.00
23) Nitrobenzene-d5	7.24	82	482395	73.41	ng	-0.01
42) 2,4,6-Tribromophenol	10.50	330	180312	91.18	ng	0.00
45) 2-Fluorobiphenyl	9.03	172	882915	86.34	ng	0.00
79) Terphenyl-d14	12.77	244	730992	55.10	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.43	163	183041	15.084	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF041619\
Data File : BF113794.D
Acq On : 16 Apr 2019 15:27
Operator : JU/SJ
Sample : K2422-03
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP1-S

Quant Time: Apr 16 18:08:15 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF040319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 16 17:22:45 2019
Response via : Initial Calibration

