

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF041819\
 Data File : BF113818.D
 Acq On : 18 Apr 2019 15:34
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 18 18:26:14 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 18 13:55:51 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.599	0.603	-0.7	103	0.00
3	Pyridine	1.647	1.671	-1.5	104	0.00
4	n-Nitrosodimethylamine	0.575	0.577	-0.3	102	0.00
5 S	2-Fluorophenol	1.203	1.230	-2.2	102	0.00
6	Aniline	2.129	2.141	-0.6	101	0.00
7 S	Phenol-d6	1.500	1.528	-1.9	102	0.00
8	2-Chlorophenol	1.350	1.386	-2.7	102	0.00
9	Benzaldehyde	0.892	0.904	-1.3	105	0.00
10 C	Phenol	1.681	1.725	-2.6	103	0.00
11	bis(2-Chloroethyl)ether	1.365	1.386	-1.5	102	0.00
12	1,3-Dichlorobenzene	1.500	1.546	-3.1	103	0.00
13 C	1,4-Dichlorobenzene	1.501	1.543	-2.8	103	0.00
14	1,2-Dichlorobenzene	1.383	1.445	-4.5	104	0.00
15	Benzyl Alcohol	1.215	1.258	-3.5	103	0.00
16	2,2'-oxybis(1-Chloropropane	1.656	1.672	-1.0	102	0.00
17	2-Methylphenol	1.103	1.110	-0.6	102	0.00
18	Hexachloroethane	0.538	0.557	-3.5	104	0.00
19 P	n-Nitroso-di-n-propylamine	1.001	1.003	-0.2	102	0.00
20	3+4-Methylphenols	1.325	1.377	-3.9	103	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
22	Acetophenone	0.456	0.464	-1.8	102	0.00
23 S	Nitrobenzene-d5	0.328	0.344	-4.9	105	0.00
24	Nitrobenzene	0.369	0.381	-3.3	103	0.00
25	Isophorone	0.680	0.676	0.6	102	0.00
26 C	2-Nitrophenol	0.147	0.162	-10.2	112	0.00
27	2,4-Dimethylphenol	0.280	0.285	-1.8	102	0.00
28	bis(2-Chloroethoxy)methane	0.432	0.435	-0.7	102	0.00
29 C	2,4-Dichlorophenol	0.298	0.304	-2.0	103	0.00
30	1,2,4-Trichlorobenzene	0.331	0.339	-2.4	103	0.00
31	Naphthalene	0.946	0.978	-3.4	103	0.00
32	Benzoic acid	0.173	0.188	-8.7	112	0.00
33	4-Chloroaniline	0.401	0.403	-0.5	102	0.00
34 C	Hexachlorobutadiene	0.202	0.210	-4.0	104	0.00
35	Caprolactam	0.096	0.097	-1.0	102	0.00
36 C	4-Chloro-3-methylphenol	0.296	0.299	-1.0	102	0.00
37	2-Methylnaphthalene	0.628	0.640	-1.9	102	0.00
38	1-Methylnaphthalene	0.609	0.623	-2.3	102	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	0.635	0.651	-2.5	102	0.00
41 P	Hexachlorocyclopentadiene	0.349	0.352	-0.9	100	0.00
42 S	2,4,6-Tribromophenol	0.221	0.232	-5.0	104	0.00
43 C	2,4,6-Trichlorophenol	0.409	0.417	-2.0	103	0.00
44	2,4,5-Trichlorophenol	0.441	0.457	-3.6	103	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF041819\
 Data File : BF113818.D
 Acq On : 18 Apr 2019 15:34
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 18 18:26:14 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 18 13:55:51 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.311	1.258	4.0	101	0.00
46	1,1'-Biphenyl	1.513	1.569	-3.7	102	0.00
47	2-Chloronaphthalene	1.239	1.269	-2.4	102	0.00
48	2-Nitroaniline	0.322	0.353	-9.6	108	0.00
49	Acenaphthylene	1.938	1.995	-2.9	102	0.00
50	Dimethylphthalate	1.400	1.427	-1.9	101	0.00
51	2,6-Dinitrotoluene	0.309	0.327	-5.8	106	0.00
52 C	Acenaphthene	1.137	1.169	-2.8	103	0.00
53	3-Nitroaniline	0.327	0.348	-6.4	106	0.00
54 P	2,4-Dinitrophenol	0.094	0.104	-10.6	117	0.00
55	Dibenzofuran	1.703	1.757	-3.2	102	0.00
56 P	4-Nitrophenol	0.266	0.286	-7.5	105	0.00
57	2,4-Dinitrotoluene	0.383	0.418	-9.1	108	0.00
58	Fluorene	1.270	1.287	-1.3	101	0.00
59	2,3,4,6-Tetrachlorophenol	0.376	0.390	-3.7	103	0.00
60	Diethylphthalate	1.341	1.363	-1.6	102	0.00
61	4-Chlorophenyl-phenylether	0.650	0.675	-3.8	102	0.00
62	4-Nitroaniline	0.315	0.324	-2.9	104	0.00
63	Azobenzene	1.378	1.409	-2.2	101	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	0.076	0.084	-10.5	110	0.00
66 c	n-Nitrosodiphenylamine	0.606	0.619	-2.1	102	0.00
67	4-Bromophenyl-phenylether	0.235	0.239	-1.7	100	0.00
68	Hexachlorobenzene	0.258	0.262	-1.6	100	0.00
69	Atrazine	0.188	0.195	-3.7	101	0.00
70 C	Pentachlorophenol	0.146	0.154	-5.5	105	0.00
71	Phenanthrene	1.019	1.042	-2.3	101	0.00
72	Anthracene	1.026	1.053	-2.6	101	0.00
73	Carbazole	0.932	0.949	-1.8	100	0.00
74	Di-n-butylphthalate	1.037	1.074	-3.6	102	0.00
75 C	Fluoranthene	1.080	1.100	-1.9	99	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
77	Benzidine	0.579	0.722	-24.7	109	0.00
78	Pyrene	1.410	1.429	-1.3	99	0.00
79 S	Terphenyl-d14	0.954	0.975	-2.2	100	0.00
80	Butylbenzylphthalate	0.522	0.535	-2.5	100	0.00
81	Benzo(a)anthracene	1.169	1.173	-0.3	96	0.00
82	3,3'-Dichlorobenzidine	0.416	0.441	-6.0	97	0.00
83	Chrysene	1.160	1.171	-0.9	97	0.00
84	Bis(2-ethylhexyl)phthalate	0.625	0.636	-1.8	99	0.00
85 c	Di-n-octyl phthalate	0.969	0.961	0.8	95	0.00
86	Indeno(1,2,3-cd)pyrene	0.947	0.889	6.1	101	0.00
87 I	Perylene-d12	1.000	1.000	0.0	93	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF041819\
Data File : BF113818.D
Acq On : 18 Apr 2019 15:34
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 18 18:26:14 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF041819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 18 13:55:51 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.247	1.336	-7.1	96	0.00
89	Benzo(k)fluoranthene	1.136	1.139	-0.3	90	0.00
90 C	Benzo(a)pyrene	1.102	1.118	-1.5	94	0.00
91	Dibenzo(a,h)anthracene	0.981	0.984	-0.3	100	0.00
92	Benzo(q,h,i)perylene	0.997	1.002	-0.5	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0