

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042123\
 Data File : BF132973.D
 Acq On : 21 Apr 2023 17:22
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF042123

Quant Time: Apr 22 04:10:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Apr 22 03:30:56 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	99	0.00
2	1,4-Dioxane	40.000	40.569	-1.4	100	0.00
3	Pyridine	40.000	42.338	-5.8	102	0.00
4	n-Nitrosodimethylamine	40.000	40.811	-2.0	100	0.00
5 S	2-Fluorophenol	80.000	79.399	0.8	98	0.00
6	Aniline	40.000	42.396	-6.0	101	0.00
7 S	Phenol-d6	80.000	80.905	-1.1	101	0.00
8	2-Chlorophenol	40.000	40.656	-1.6	100	0.00
9	Benzaldehyde	40.000	43.130	-7.8	103	0.00
10 C	Phenol	40.000	40.956	-2.4	101	0.00
11	bis(2-Chloroethyl)ether	40.000	40.460	-1.2	99	0.00
12	1,3-Dichlorobenzene	40.000	39.974	0.1	99	0.00
13 C	1,4-Dichlorobenzene	40.000	40.400	-1.0	98	0.00
14	1,2-Dichlorobenzene	40.000	40.787	-2.0	100	0.00
15	Benzyl Alcohol	40.000	41.910	-4.8	99	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.600	-1.5	99	0.00
17	2-Methylphenol	40.000	40.559	-1.4	99	0.00
18	Hexachloroethane	40.000	40.812	-2.0	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.006	-0.0	101	0.00
20	3+4-Methylphenols	40.000	40.322	-0.8	100	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
22	Acetophenone	40.000	39.044	2.4	100	0.00
23 S	Nitrobenzene-d5	80.000	82.178	-2.7	101	0.00
24	Nitrobenzene	40.000	40.915	-2.3	101	0.00
25	Isophorone	40.000	40.711	-1.8	101	0.00
26 C	2-Nitrophenol	40.000	43.225	-8.1	102	0.00
27	2,4-Dimethylphenol	40.000	40.982	-2.5	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.274	-0.7	100	0.00
29 C	2,4-Dichlorophenol	40.000	40.827	-2.1	99	0.00
30	1,2,4-Trichlorobenzene	40.000	40.084	-0.2	100	0.00
31	Naphthalene	40.000	40.276	-0.7	100	0.00
32	Benzoic acid	40.000	47.291	-18.2	108	0.00
33	4-Chloroaniline	40.000	40.328	-0.8	101	0.00
34 C	Hexachlorobutadiene	40.000	39.654	0.9	98	0.00
35	Caprolactam	40.000	42.759	-6.9	102	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.352	-3.4	101	0.00
37	2-Methylnaphthalene	40.000	39.806	0.5	99	0.00
38	1-Methylnaphthalene	40.000	40.180	-0.4	100	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.192	-0.5	100	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.641	-6.6	101	0.00
42 S	2,4,6-Tribromophenol	80.000	83.016	-3.8	101	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.099	-2.7	101	0.00
44	2,4,5-Trichlorophenol	40.000	40.369	-0.9	98	0.00
45 S	2-Fluorobiphenyl	80.000	82.330	-2.9	102	0.00
46	1,1'-Biphenyl	40.000	39.641	0.9	99	0.00
47	2-Chloronaphthalene	40.000	40.022	-0.1	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042123\
 Data File : BF132973.D
 Acq On : 21 Apr 2023 17:22
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF042123

Quant Time: Apr 22 04:10:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Apr 22 03:30:56 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.775	-6.9	103	0.00
49	Acenaphthylene	40.000	40.363	-0.9	99	0.00
50	Dimethylphthalate	40.000	40.288	-0.7	100	0.00
51	2,6-Dinitrotoluene	40.000	41.634	-4.1	102	0.00
52 C	Acenaphthene	40.000	40.070	-0.2	99	0.00
53	3-Nitroaniline	40.000	41.842	-4.6	100	0.00
54 P	2,4-Dinitrophenol	40.000	45.935	-14.8	104	0.00
55	Dibenzofuran	40.000	40.001	-0.0	100	0.00
56 P	4-Nitrophenol	40.000	41.955	-4.9	99	0.00
57	2,4-Dinitrotoluene	40.000	42.537	-6.3	101	0.00
58	Fluorene	40.000	38.349	4.1	99	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.733	-1.8	101	0.00
60	Diethylphthalate	40.000	40.735	-1.8	101	0.00
61	4-Chlorophenyl-phenylether	40.000	38.914	2.7	99	0.00
62	4-Nitroaniline	40.000	41.767	-4.4	104	0.00
63	Azobenzene	40.000	40.184	-0.5	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.156	-15.4	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.690	-1.7	101	0.00
67	4-Bromophenyl-phenylether	40.000	41.259	-3.1	101	0.00
68	Hexachlorobenzene	40.000	40.823	-2.1	100	0.00
69	Atrazine	40.000	41.613	-4.0	99	0.00
70 C	Pentachlorophenol	40.000	44.245	-10.6	102	0.00
71	Phenanthrene	40.000	39.691	0.8	99	0.00
72	Anthracene	40.000	40.484	-1.2	100	0.00
73	Carbazole	40.000	40.780	-2.0	100	0.00
74	Di-n-butylphthalate	40.000	42.036	-5.1	100	0.00
75 C	Fluoranthene	40.000	41.055	-2.6	100	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzydine	40.000	34.249	14.4	86	0.00
78	Pyrene	40.000	39.982	0.0	99	0.00
79 S	Terphenyl-d14	80.000	78.973	1.3	99	0.00
80	Butylbenzylphthalate	40.000	43.957	-9.9	102	0.00
81	Benzo(a)anthracene	40.000	40.848	-2.1	101	0.00
82	3,3'-Dichlorobenzidine	40.000	41.459	-3.6	97	0.00
83	Chrysene	40.000	40.150	-0.4	99	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.495	-6.2	101	0.00
85 c	Di-n-octyl phthalate	40.000	43.671	-9.2	102	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.042	-7.6	101	0.00
88	Benzo(b)fluoranthene	40.000	43.700	-9.3	104	0.00
89	Benzo(k)fluoranthene	40.000	39.039	2.4	95	0.00
90 C	Benzo(a)pyrene	40.000	41.888	-4.7	101	0.00
91	Dibenzo(a,h)anthracene	40.000	43.314	-8.3	100	0.00
92	Benzo(g,h,i)perylene	40.000	43.005	-7.5	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042123\
Data File : BF132973.D
Acq On : 21 Apr 2023 17:22
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF042123

Quant Time: Apr 22 04:10:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Apr 22 03:30:56 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0