

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042123\
 Data File : BF132973.D
 Acq On : 21 Apr 2023 17:22
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF042123

Quant Time: Apr 22 04:10:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Apr 22 03:30:56 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.614	0.623	-1.5	100	0.00
3	Pyridine	1.631	1.726	-5.8	102	0.00
4	n-Nitrosodimethylamine	0.845	0.862	-2.0	100	0.00
5 S	2-Fluorophenol	1.324	1.314	0.8	98	0.00
6	Aniline	2.117	2.243	-6.0	101	0.00
7 S	Phenol-d6	1.643	1.662	-1.2	101	0.00
8	2-Chlorophenol	1.344	1.366	-1.6	100	0.00
9	Benzaldehyde	0.791	0.795	-0.5	103	0.00
10 C	Phenol	1.813	1.856	-2.4	101	0.00
11	bis(2-Chloroethyl)ether	1.360	1.376	-1.2	99	0.00
12	1,3-Dichlorobenzene	1.429	1.428	0.1	99	0.00
13 C	1,4-Dichlorobenzene	1.432	1.447	-1.0	98	0.00
14	1,2-Dichlorobenzene	1.321	1.347	-2.0	100	0.00
15	Benzyl Alcohol	1.135	1.189	-4.8	99	0.00
16	2,2'-oxybis(1-Chloropropane	2.363	2.398	-1.5	99	0.00
17	2-Methylphenol	1.231	1.248	-1.4	99	0.00
18	Hexachloroethane	0.498	0.508	-2.0	100	0.00
19 P	n-Nitroso-di-n-propylamine	1.023	1.023	0.0	101	0.00
20	3+4-Methylphenols	1.450	1.461	-0.8	100	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
22	Acetophenone	0.463	0.452	2.4	100	0.00
23 S	Nitrobenzene-d5	0.371	0.381	-2.7	101	0.00
24	Nitrobenzene	0.375	0.384	-2.4	101	0.00
25	Isophorone	0.704	0.716	-1.7	101	0.00
26 C	2-Nitrophenol	0.181	0.196	-8.3	102	0.00
27	2,4-Dimethylphenol	0.311	0.318	-2.3	100	0.00
28	bis(2-Chloroethoxy)methane	0.416	0.419	-0.7	100	0.00
29 C	2,4-Dichlorophenol	0.283	0.289	-2.1	99	0.00
30	1,2,4-Trichlorobenzene	0.293	0.293	0.0	100	0.00
31	Naphthalene	1.000	1.007	-0.7	100	0.00
32	Benzoic acid	0.192	0.227	-18.2	108	0.00
33	4-Chloroaniline	0.404	0.407	-0.7	101	0.00
34 C	Hexachlorobutadiene	0.162	0.161	0.6	98	0.00
35	Caprolactam	0.095	0.102	-7.4	102	0.00
36 C	4-Chloro-3-methylphenol	0.305	0.315	-3.3	101	0.00
37	2-Methylnaphthalene	0.684	0.680	0.6	99	0.00
38	1-Methylnaphthalene	0.640	0.642	-0.3	100	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	0.538	0.540	-0.4	100	0.00
41 P	Hexachlorocyclopentadiene	0.314	0.334	-6.4	101	0.00
42 S	2,4,6-Tribromophenol	0.201	0.209	-4.0	101	0.00
43 C	2,4,6-Trichlorophenol	0.372	0.382	-2.7	101	0.00
44	2,4,5-Trichlorophenol	0.430	0.434	-0.9	98	0.00
45 S	2-Fluorobiphenyl	1.195	1.230	-2.9	102	0.00
46	1,1'-Biphenyl	1.586	1.572	0.9	99	0.00
47	2-Chloronaphthalene	1.161	1.161	0.0	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042123\
 Data File : BF132973.D
 Acq On : 21 Apr 2023 17:22
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF042123

Quant Time: Apr 22 04:10:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Apr 22 03:30:56 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.384	0.410	-6.8	103	0.00
49	Acenaphthylene	1.855	1.872	-0.9	99	0.00
50	Dimethylphthalate	1.395	1.405	-0.7	100	0.00
51	2,6-Dinitrotoluene	0.314	0.327	-4.1	102	0.00
52 C	Acenaphthene	1.242	1.244	-0.2	99	0.00
53	3-Nitroaniline	0.373	0.390	-4.6	100	0.00
54 P	2,4-Dinitrophenol	0.158	0.181	-14.6	104	0.00
55	Dibenzofuran	1.695	1.695	0.0	100	0.00
56 P	4-Nitrophenol	0.289	0.303	-4.8	99	0.00
57	2,4-Dinitrotoluene	0.400	0.426	-6.5	101	0.00
58	Fluorene	1.241	1.190	4.1	99	0.00
59	2,3,4,6-Tetrachlorophenol	0.333	0.340	-2.1	101	0.00
60	Diethylphthalate	1.317	1.341	-1.8	101	0.00
61	4-Chlorophenyl-phenylether	0.605	0.589	2.6	99	0.00
62	4-Nitroaniline	0.343	0.358	-4.4	104	0.00
63	Azobenzene	1.395	1.402	-0.5	100	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	0.121	0.140	-15.7	103	0.00
66 c	n-Nitrosodiphenylamine	0.649	0.660	-1.7	101	0.00
67	4-Bromophenyl-phenylether	0.217	0.224	-3.2	101	0.00
68	Hexachlorobenzene	0.222	0.227	-2.3	100	0.00
69	Atrazine	0.187	0.194	-3.7	99	0.00
70 C	Pentachlorophenol	0.158	0.175	-10.8	102	0.00
71	Phenanthrene	1.075	1.067	0.7	99	0.00
72	Anthracene	1.100	1.113	-1.2	100	0.00
73	Carbazole	0.980	0.999	-1.9	100	0.00
74	Di-n-butylphthalate	1.109	1.165	-5.0	100	0.00
75 C	Fluoranthene	1.079	1.107	-2.6	100	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
77	Benzydine	0.338	0.290	14.2	86	0.00
78	Pyrene	1.714	1.714	0.0	99	0.00
79 S	Terphenyl-d14	1.160	1.145	1.3	99	0.00
80	Butylbenzylphthalate	0.607	0.667	-9.9	102	0.00
81	Benzo(a)anthracene	1.375	1.404	-2.1	101	0.00
82	3,3'-Dichlorobenzidine	0.380	0.394	-3.7	97	0.00
83	Chrysene	1.375	1.380	-0.4	99	0.00
84	Bis(2-ethylhexyl)phthalate	0.762	0.809	-6.2	101	0.00
85 c	Di-n-octyl phthalate	1.242	1.356	-9.2	102	0.00
86 I	Perylene-d12	1.000	1.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	1.138	1.224	-7.6	101	0.00
88	Benzo(b)fluoranthene	1.272	1.390	-9.3	104	0.00
89	Benzo(k)fluoranthene	1.261	1.231	2.4	95	0.00
90 C	Benzo(a)pyrene	1.158	1.213	-4.7	101	0.00
91	Dibenzo(a,h)anthracene	0.949	1.027	-8.2	100	0.00
92	Benzo(g,h,i)perylene	0.937	1.007	-7.5	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042123\
Data File : BF132973.D
Acq On : 21 Apr 2023 17:22
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF042123

Quant Time: Apr 22 04:10:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Apr 22 03:30:56 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range SPCC's out = 0 CCC's out = 0