

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042125\
 Data File : BF142199.D
 Acq On : 21 Apr 2025 11:13
 Operator : RC/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Quant Time: Apr 22 00:46:42 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 22 00:36:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.863	152	622834	20.000	ng	0.00
21) Naphthalene-d8	8.145	136	2413530	20.000	ng	0.00
39) Acenaphthene-d10	9.898	164	1431118	20.000	ng	0.00
64) Phenanthrene-d10	11.386	188	2538453	20.000	ng	0.00
76) Chrysene-d12	14.027	240	1890280	20.000	ng	0.00
86) Perylene-d12	15.504	264	1397277	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042125\
Data File : BF142199.D
Acq On : 21 Apr 2025 11:13
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Quant Time: Apr 22 00:46:42 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042125.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 22 00:36:52 2025
Response via : Initial Calibration

