





Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042219.M
Calibration Table Last Updated: Tue Apr 23 03:32:35 2019

2,4-Dinitrophenol


Response Ratio



Resp Ratio = $1.50 \times 10^{-1} \times \text{Amt} - 6.23 \times 10^{-2}$
Coef of Det (r^2) = 0.996 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042219.M
Calibration Table Last Updated: Tue Apr 23 03:32:35 2019

4,6-Dinitro-2-methylphenol



Resp Ratio = $1.08 \times 10^{-1} \times \text{Amt} - 3.95 \times 10^{-2}$
Coef of Det (r^2) = 0.998 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042219.M
Calibration Table Last Updated: Tue Apr 23 03:32:35 2019