




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042219\
Data File : BF113874.D
Acq On : 22 Apr 2019 12:20
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 29 05:11:34 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 23 03:32:35 2019
Response via : Initial Calibration



TIC: BF113874.D

(70) Pentachlorophenol (C)




11.222min (+0.000) 38457.79ng

response 313916

Ion	Exp%	Act%
265.70	100	100
268.00	66.90	64.58
264.00	63.80	64.21
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042219\
 Data File : BF113874.D
 Acq On : 22 Apr 2019 12:20
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 29 05:11:34 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 23 03:32:35 2019
 Response via : Initial Calibration



TIC: BF113874.D

(77) Benzidine

12.751min (+0.007) 0.00ng

response 1851542

Ion	Exp%	Act%
184.00	100	100
185.00	16.30	14.55
183.00	11.70	12.27
0.00	0.00	0.00