

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042222\
 Data File : BF127865.D
 Acq On : 22 Apr 2022 22:26
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF042222

Quant Time: Apr 24 00:50:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Apr 24 00:43:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	40.847	-2.1	103	0.00
3	Pyridine	40.000	42.401	-6.0	104	0.00
4	n-Nitrosodimethylamine	40.000	41.084	-2.7	103	0.00
5 S	2-Fluorophenol	80.000	80.613	-0.8	105	0.00
6	Aniline	40.000	42.430	-6.1	105	0.00
7 S	Phenol-d6	80.000	81.681	-2.1	104	0.00
8	2-Chlorophenol	40.000	40.937	-2.3	104	0.00
9	Benzaldehyde	40.000	36.000	10.0	105	0.00
10 C	Phenol	40.000	40.985	-2.5	104	0.00
11	bis(2-Chloroethyl)ether	40.000	41.098	-2.7	105	0.00
12	1,3-Dichlorobenzene	40.000	40.267	-0.7	104	0.00
13 C	1,4-Dichlorobenzene	40.000	40.291	-0.7	104	0.00
14	1,2-Dichlorobenzene	40.000	39.980	0.1	104	0.00
15	Benzyl Alcohol	40.000	44.493	-11.2	107	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.680	-4.2	104	0.00
17	2-Methylphenol	40.000	41.938	-4.8	103	0.00
18	Hexachloroethane	40.000	40.510	-1.3	104	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.661	-1.7	104	0.00
20	3+4-Methylphenols	40.000	40.901	-2.3	103	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	104	0.00
22	Acetophenone	40.000	38.888	2.8	103	0.00
23 S	Nitrobenzene-d5	80.000	80.928	-1.2	104	0.00
24	Nitrobenzene	40.000	41.018	-2.5	105	0.00
25	Isophorone	40.000	40.376	-0.9	103	0.00
26 C	2-Nitrophenol	40.000	43.609	-9.0	107	0.00
27	2,4-Dimethylphenol	40.000	41.512	-3.8	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.475	-1.2	104	0.00
29 C	2,4-Dichlorophenol	40.000	40.671	-1.7	104	0.00
30	1,2,4-Trichlorobenzene	40.000	39.400	1.5	103	0.00
31	Naphthalene	40.000	39.185	2.0	102	0.00
32	Benzoic acid	40.000	45.634	-14.1	107	0.00
33	4-Chloroaniline	40.000	40.549	-1.4	103	0.00
34 C	Hexachlorobutadiene	40.000	39.905	0.2	104	0.00
35	Caprolactam	40.000	41.177	-2.9	102	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.375	-3.4	105	0.00
37	2-Methylnaphthalene	40.000	39.544	1.1	103	0.00
38	1-Methylnaphthalene	40.000	39.439	1.4	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.617	1.0	102	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.800	-7.0	104	0.00
42 S	2,4,6-Tribromophenol	80.000	80.078	-0.1	103	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.176	-2.9	105	0.00
44	2,4,5-Trichlorophenol	40.000	40.403	-1.0	101	0.00
45 S	2-Fluorobiphenyl	80.000	80.140	-0.2	102	0.00
46	1,1'-Biphenyl	40.000	39.453	1.4	102	0.00
47	2-Chloronaphthalene	40.000	39.586	1.0	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042222\
 Data File : BF127865.D
 Acq On : 22 Apr 2022 22:26
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF042222

Quant Time: Apr 24 00:50:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Apr 24 00:43:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.233	-5.6	104	0.00
49	Acenaphthylene	40.000	40.164	-0.4	103	0.00
50	Dimethylphthalate	40.000	40.363	-0.9	104	0.00
51	2,6-Dinitrotoluene	40.000	41.734	-4.3	103	0.00
52 C	Acenaphthene	40.000	40.396	-1.0	103	0.00
53	3-Nitroaniline	40.000	41.080	-2.7	101	0.00
54 P	2,4-Dinitrophenol	40.000	45.665	-14.2	108	0.00
55	Dibenzofuran	40.000	40.025	-0.1	103	0.00
56 P	4-Nitrophenol	40.000	42.487	-6.2	102	0.00
57	2,4-Dinitrotoluene	40.000	42.175	-5.4	103	0.00
58	Fluorene	40.000	38.481	3.8	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.627	-4.1	104	0.00
60	Diethylphthalate	40.000	39.865	0.3	101	0.00
61	4-Chlorophenyl-phenylether	40.000	38.524	3.7	101	0.00
62	4-Nitroaniline	40.000	40.933	-2.3	100	0.00
63	Azobenzene	40.000	40.135	-0.3	102	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.397	-16.0	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.102	-0.3	102	0.00
67	4-Bromophenyl-phenylether	40.000	40.952	-2.4	103	0.00
68	Hexachlorobenzene	40.000	39.879	0.3	101	0.00
69	Atrazine	40.000	39.385	1.5	101	0.00
70 C	Pentachlorophenol	40.000	42.530	-6.3	100	0.00
71	Phenanthrene	40.000	39.427	1.4	100	0.00
72	Anthracene	40.000	39.779	0.6	101	0.00
73	Carbazole	40.000	38.950	2.6	99	0.00
74	Di-n-butylphthalate	40.000	40.160	-0.4	101	0.00
75 C	Fluoranthene	40.000	39.037	2.4	100	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	97	0.00
77	Benzydine	40.000	47.338	-18.3	107	0.00
78	Pyrene	40.000	40.760	-1.9	99	0.00
79 S	Terphenyl-d14	80.000	78.499	1.9	99	0.00
80	Butylbenzylphthalate	40.000	42.785	-7.0	101	0.00
81	Benzo(a)anthracene	40.000	41.220	-3.0	99	0.00
82	3,3'-Dichlorobenzidine	40.000	39.847	0.4	97	0.00
83	Chrysene	40.000	40.377	-0.9	99	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.162	-5.4	100	0.00
85 c	Di-n-octyl phthalate	40.000	41.948	-4.9	98	0.00
86 I	Perylene-d12	20.000	20.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.425	-8.6	102	0.00
88	Benzo(b)fluoranthene	40.000	40.011	-0.0	99	0.00
89	Benzo(k)fluoranthene	40.000	40.267	-0.7	96	0.00
90 C	Benzo(a)pyrene	40.000	40.976	-2.4	98	0.00
91	Dibenzo(a,h)anthracene	40.000	43.909	-9.8	102	0.00
92	Benzo(g,h,i)perylene	40.000	44.786	-12.0	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042222\
Data File : BF127865.D
Acq On : 22 Apr 2022 22:26
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF042222

Quant Time: Apr 24 00:50:35 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Apr 24 00:43:03 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0