



Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF042222\  
 Data File : BF127856.D  
 Acq On : 22 Apr 2022 17:41  
 Operator : CG\JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF042222.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Sun Apr 24 00:43:03 2022






AutoFind: Scans 1602, 1603, 1604; Background Corrected with Scan 1596

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 42.9         | 104528     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.5          | 1571       | PASS                |
| 69             | 198             | 0.00            | 100             | 42.5         | 103496     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.5          | 491        | PASS                |
| 127            | 198             | 10              | 80              | 48.9         | 119072     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.7          | 1796       | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 243744     | PASS                |
| 199            | 198             | 5               | 9               | 6.7          | 16312      | PASS                |
| 275            | 198             | 10              | 60              | 26.1         | 63603      | PASS                |
| 365            | 198             | 1               | 100             | 3.5          | 8590       | PASS                |
| 441            | 198             | 0.01            | 100             | 13.3         | 32317      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 201976     | PASS                |
| 443            | 442             | 15              | 24              | 19.7         | 39724      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF042222\  
Data File : BF127856.D  
Acq On : 22 Apr 2022 17:41  
Operator : CG\JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Quant Time: Apr 24 04:50:28 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF042222.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Sun Apr 24 00:43:03 2022  
Response via : Initial Calibration



TIC: BF127856.D\data.ms

**(70) Pentachlorophenol (C)**

**11.275min (+ 0.000) 624957.94 ng**




**response 236180**

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 66.30  | 67.00  |
| 264.00 | 66.40  | 63.78  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF042222\  
 Data File : BF127856.D  
 Acq On : 22 Apr 2022 17:41  
 Operator : CG\JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Quant Time: Apr 24 04:50:28 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF042222.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Sun Apr 24 00:43:03 2022  
 Response via : Initial Calibration



TIC: BF127856.D\data.ms

**(77) Benzidine**

12.816min (-12.816) 0.00 ng

response 0

| Ion    | Exp%   | Act%  |
|--------|--------|-------|
| 184.00 | 100.00 | 0.00  |
| 185.00 | 15.10  | 0.00# |
| 183.00 | 11.20  | 0.00# |
| 0.00   | 0.00   | 0.00  |

DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 4/22/2022     | BNA_F            | <u>BF127856.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 583403           | 13.639             |
| DDD           | 7774             | 13.333             |
| DDE           | 1015             | 13.004             |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 8789          | 592192           | 1.48               |
|               |                  |                    |