

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF042320\
 Data File : BF120184.D
 Acq On : 24 Apr 2020 2:51
 Operator : MA/SJ
 Sample : L2341-03
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 BUS-WASH

Manual Integrations
 APPROVED

mohammad
 4/24/2020 1:34:43 PM

Quant Time: Apr 24 04:58:01 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF040820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 23 13:31:01 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.66	152	136828	20.00	ng	0.00
21) Naphthalene-d8	7.95	136	509962	20.00	ng	0.00
39) Acenaphthene-d10	9.70	164	275632	20.00	ng	0.00
64) Phenanthrene-d10	11.18	188	506057	20.00	ng	0.00
76) Chrysene-d12	13.82	240	330961	20.00	ng	0.00
87) Perylene-d12	15.22	264	269535	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.27	112	907446	114.61	ng	0.01
7) Phenol-d6	6.31	99	1144074	112.14	ng	0.00
23) Nitrobenzene-d5	7.23	82	890744	90.36	ng	0.00
42) 2,4,6-Tribromophenol	10.49	330	398538	118.10	ng	0.00
45) 2-Fluorobiphenyl	9.03	172	1786499	92.02	ng	0.00
79) Terphenyl-d14	12.78	244	1813330	92.19	ng	0.00
Target Compounds						
10) Phenol	6.32	94	23574	2.04	ng	Qvalue 89
20) 3+4-Methylphenols	7.09	107	241914	25.05	ng	# 75
32) Benzoic acid	8.05	122	5630337m	858.01	ng	
50) Dimethylphthalate	9.43	163	138065	6.48	ng	99
84) Bis(2-ethylhexyl)phthalate	13.82	149	51187	2.79	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF042320\
Data File : BF120184.D
Acq On : 24 Apr 2020 2:51
Operator : MA/SJ
Sample : L2341-03
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
BUS-WASH

Manual Integrations
APPROVED

mohammad
4/24/2020 1:34:43 PM

Quant Time: Apr 24 04:58:01 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF040820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 23 13:31:01 2020
Response via : Initial Calibration

