

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF042420\
 Data File : BF120194.D
 Acq On : 24 Apr 2020 11:34
 Operator : MA/SJ
 Sample : L2360-01
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 GREENWALL

Quant Time: Apr 24 12:40:37 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF040820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 24 10:24:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.66	152	142444	20.00	ng	0.00
21) Naphthalene-d8	7.95	136	543977	20.00	ng	0.00
39) Acenaphthene-d10	9.70	164	278120	20.00	ng	0.00
64) Phenanthrene-d10	11.18	188	480648	20.00	ng	0.00
76) Chrysene-d12	13.82	240	281376	20.00	ng	-0.01
87) Perylene-d12	15.22	264	254725	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.27	112	824655	100.05	ng	0.00
7) Phenol-d6	6.31	99	1122384	105.68	ng	0.00
23) Nitrobenzene-d5	7.23	82	736146	70.01	ng	0.00
42) 2,4,6-Tribromophenol	10.49	330	254035	74.60	ng	0.00
45) 2-Fluorobiphenyl	9.03	172	1415753	72.27	ng	0.00
79) Terphenyl-d14	12.77	244	1267376	75.79	ng	0.00
Target Compounds						
31) Naphthalene	7.97	128	110017	3.844	ng	96
50) Dimethylphthalate	9.42	163	226618	10.534	ng	99
66) n-Nitrosodiphenylamine	10.36	169	182527	11.419	ng	97
74) Di-n-butylphthalate	11.76	149	156344	4.813	ng	98
84) Bis(2-ethylhexyl)phthalate	13.81	149	111139	7.130	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF042420\
Data File : BF120194.D
Acq On : 24 Apr 2020 11:34
Operator : MA/SJ
Sample : L2360-01
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
GREENWALL

Quant Time: Apr 24 12:40:37 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF040820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 24 10:24:30 2020
Response via : Initial Calibration

