




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042424\
Data File : BF137708.D
Acq On : 24 Apr 2024 09:26
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 24 13:59:48 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 24 03:57:12 2024
Response via : Initial Calibration



TIC: BF137708.D\data.ms

(70) Pentachlorophenol (C)




11.398min (+ 0.000) 304847.69 ng

response 512559

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	64.21
264.00	64.20	64.14
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042424\
Data File : BF137708.D
Acq On : 24 Apr 2024 09:26
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 24 13:59:48 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 24 03:57:12 2024
Response via : Initial Calibration



TIC: BF137708.D\data.ms

(77) Benzidine

12.945min (+ 0.006) 174639.08 ng

response 2537134

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.30	15.47
183.00	11.70	12.16
0.00	0.00	0.00