

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042424\
 Data File : BF137709.D
 Acq On : 24 Apr 2024 09:56
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 24 10:27:17 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 24 03:57:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	91	0.00
2	1,4-Dioxane	40.000	38.323	4.2	90	-0.01
3	Pyridine	40.000	38.195	4.5	89	0.00
4	n-Nitrosodimethylamine	40.000	37.161	7.1	87	-0.01
5 S	2-Fluorophenol	80.000	75.403	5.7	90	0.00
6	Aniline	40.000	37.810	5.5	89	0.00
7 S	Phenol-d6	80.000	75.859	5.2	89	0.00
8	2-Chlorophenol	40.000	38.670	3.3	91	0.00
9	Benzaldehyde	40.000	34.658	13.4	88	0.00
10 C	Phenol	40.000	37.798	5.5	89	0.00
11	bis(2-Chloroethyl)ether	40.000	37.782	5.5	89	0.00
12	1,3-Dichlorobenzene	40.000	38.417	4.0	91	0.00
13 C	1,4-Dichlorobenzene	40.000	38.232	4.4	90	0.00
14	1,2-Dichlorobenzene	40.000	38.434	3.9	91	0.00
15	Benzyl Alcohol	40.000	38.341	4.1	89	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.194	7.0	87	0.00
17	2-Methylphenol	40.000	37.813	5.5	89	0.00
18	Hexachloroethane	40.000	38.379	4.1	90	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.467	6.3	90	0.00
20	3+4-Methylphenols	40.000	38.398	4.0	89	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	89	0.00
22	Acetophenone	40.000	37.856	5.4	89	0.00
23 S	Nitrobenzene-d5	80.000	77.126	3.6	90	0.00
24	Nitrobenzene	40.000	38.679	3.3	90	0.00
25	Isophorone	40.000	37.638	5.9	89	0.00
26 C	2-Nitrophenol	40.000	40.953	-2.4	92	0.00
27	2,4-Dimethylphenol	40.000	37.436	6.4	90	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.989	5.0	90	0.00
29 C	2,4-Dichlorophenol	40.000	38.873	2.8	91	0.00
30	1,2,4-Trichlorobenzene	40.000	38.416	4.0	91	0.00
31	Naphthalene	40.000	38.585	3.5	91	0.00
32	Benzoic acid	40.000	38.916	2.7	88	0.00
33	4-Chloroaniline	40.000	38.283	4.3	90	0.00
34 C	Hexachlorobutadiene	40.000	38.798	3.0	90	0.00
35	Caprolactam	40.000	38.422	3.9	90	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.671	3.3	90	0.00
37	2-Methylnaphthalene	40.000	38.393	4.0	91	0.00
38	1-Methylnaphthalene	40.000	38.534	3.7	92	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	93	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.752	5.6	92	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.199	2.0	91	0.00
42 S	2,4,6-Tribromophenol	80.000	76.934	3.8	96	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.341	4.1	92	0.00
44	2,4,5-Trichlorophenol	40.000	37.971	5.1	93	0.00
45 S	2-Fluorobiphenyl	80.000	70.910	11.4	92	0.00
46	1,1'-Biphenyl	40.000	37.479	6.3	92	0.00
47	2-Chloronaphthalene	40.000	37.194	7.0	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042424\
 Data File : BF137709.D
 Acq On : 24 Apr 2024 09:56
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 24 10:27:17 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 24 03:57:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.508	1.2	93	0.00
49	Acenaphthylene	40.000	37.603	6.0	92	0.00
50	Dimethylphthalate	40.000	37.905	5.2	95	0.00
51	2,6-Dinitrotoluene	40.000	39.323	1.7	93	0.00
52 C	Acenaphthene	40.000	37.439	6.4	93	0.00
53	3-Nitroaniline	40.000	38.938	2.7	92	0.00
54 P	2,4-Dinitrophenol	40.000	37.918	5.2	97	0.00
55	Dibenzofuran	40.000	37.623	5.9	93	0.00
56 P	4-Nitrophenol	40.000	38.580	3.6	90	0.00
57	2,4-Dinitrotoluene	40.000	40.271	-0.7	94	0.00
58	Fluorene	40.000	37.364	6.6	93	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.668	3.3	94	0.00
60	Diethylphthalate	40.000	38.159	4.6	95	0.00
61	4-Chlorophenyl-phenylether	40.000	37.916	5.2	95	0.00
62	4-Nitroaniline	40.000	39.547	1.1	93	0.00
63	Azobenzene	40.000	37.622	5.9	92	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	93	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.278	-0.7	96	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.881	5.3	94	0.00
67	4-Bromophenyl-phenylether	40.000	37.947	5.1	95	0.00
68	Hexachlorobenzene	40.000	38.583	3.5	96	0.00
69	Atrazine	40.000	39.143	2.1	96	0.00
70 C	Pentachlorophenol	40.000	39.372	1.6	95	0.00
71	Phenanthrene	40.000	37.922	5.2	95	0.00
72	Anthracene	40.000	38.145	4.6	95	0.00
73	Carbazole	40.000	37.618	6.0	93	0.00
74	Di-n-butylphthalate	40.000	38.764	3.1	97	0.00
75 C	Fluoranthene	40.000	37.914	5.2	95	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	90	0.00
77	Benzydine	40.000	46.083	-15.2	110	0.00
78	Pyrene	40.000	39.105	2.2	94	0.00
79 S	Terphenyl-d14	80.000	77.428	3.2	96	0.00
80	Butylbenzylphthalate	40.000	41.349	-3.4	96	0.00
81	Benzo(a)anthracene	40.000	38.078	4.8	90	0.00
82	3,3'-Dichlorobenzidine	40.000	37.500	6.3	89	0.00
83	Chrysene	40.000	37.902	5.2	90	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.722	-4.3	97	0.00
85 c	Di-n-octyl phthalate	40.000	39.172	2.1	88	0.00
86 I	Perylene-d12	20.000	20.000	0.0	87	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.455	-1.1	97	0.00
88	Benzo(b)fluoranthene	40.000	37.688	5.8	83	0.00
89	Benzo(k)fluoranthene	40.000	40.832	-2.1	92	0.00
90 C	Benzo(a)pyrene	40.000	38.592	3.5	86	0.00
91	Dibenzo(a,h)anthracene	40.000	40.701	-1.8	97	0.00
92	Benzo(g,h,i)perylene	40.000	41.079	-2.7	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042424\
Data File : BF137709.D
Acq On : 24 Apr 2024 09:56
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 24 10:27:17 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 24 03:57:12 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0