

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042624\
 Data File : BF137730.D
 Acq On : 26 Apr 2024 11:47
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 26 12:28:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 24 03:57:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	98	0.00
2	1,4-Dioxane	40.000	37.929	5.2	96	0.00
3	Pyridine	40.000	37.951	5.1	95	0.00
4	n-Nitrosodimethylamine	40.000	36.322	9.2	92	0.00
5 S	2-Fluorophenol	80.000	75.237	6.0	97	0.00
6	Aniline	40.000	37.082	7.3	94	0.00
7 S	Phenol-d6	80.000	74.768	6.5	95	0.00
8	2-Chlorophenol	40.000	37.874	5.3	96	0.00
9	Benzaldehyde	40.000	34.103	14.7	94	0.00
10 C	Phenol	40.000	37.906	5.2	96	0.00
11	bis(2-Chloroethyl)ether	40.000	36.400	9.0	92	0.00
12	1,3-Dichlorobenzene	40.000	37.735	5.7	96	0.00
13 C	1,4-Dichlorobenzene	40.000	37.883	5.3	97	0.00
14	1,2-Dichlorobenzene	40.000	38.103	4.7	97	0.00
15	Benzyl Alcohol	40.000	37.420	6.4	94	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.537	11.2	89	0.00
17	2-Methylphenol	40.000	37.792	5.5	96	0.00
18	Hexachloroethane	40.000	37.701	5.7	95	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.097	12.3	91	0.00
20	3+4-Methylphenols	40.000	37.476	6.3	94	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	94	0.00
22	Acetophenone	40.000	37.710	5.7	94	0.00
23 S	Nitrobenzene-d5	80.000	76.365	4.5	94	0.00
24	Nitrobenzene	40.000	38.445	3.9	94	0.00
25	Isophorone	40.000	36.602	8.5	92	0.00
26 C	2-Nitrophenol	40.000	42.005	-5.0	100	0.00
27	2,4-Dimethylphenol	40.000	37.341	6.6	95	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.002	7.5	92	0.00
29 C	2,4-Dichlorophenol	40.000	39.242	1.9	97	0.00
30	1,2,4-Trichlorobenzene	40.000	38.833	2.9	97	0.00
31	Naphthalene	40.000	38.452	3.9	96	0.00
32	Benzoic acid	40.000	41.387	-3.5	99	0.00
33	4-Chloroaniline	40.000	38.325	4.2	95	0.00
34 C	Hexachlorobutadiene	40.000	39.831	0.4	98	0.00
35	Caprolactam	40.000	37.870	5.3	93	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.149	4.6	94	0.00
37	2-Methylnaphthalene	40.000	37.975	5.1	96	0.00
38	1-Methylnaphthalene	40.000	37.665	5.8	95	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	95	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.895	2.8	97	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.551	-1.4	97	0.00
42 S	2,4,6-Tribromophenol	80.000	79.079	1.2	101	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.417	4.0	94	0.00
44	2,4,5-Trichlorophenol	40.000	38.704	3.2	97	0.00
45 S	2-Fluorobiphenyl	80.000	71.796	10.3	96	0.00
46	1,1'-Biphenyl	40.000	37.762	5.6	94	0.00
47	2-Chloronaphthalene	40.000	37.993	5.0	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042624\
 Data File : BF137730.D
 Acq On : 26 Apr 2024 11:47
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 26 12:28:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 24 03:57:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.654	0.9	95	0.00
49	Acenaphthylene	40.000	37.780	5.5	95	0.00
50	Dimethylphthalate	40.000	37.401	6.5	96	0.00
51	2,6-Dinitrotoluene	40.000	39.420	1.4	96	0.00
52 C	Acenaphthene	40.000	37.917	5.2	96	0.00
53	3-Nitroaniline	40.000	39.493	1.3	95	0.00
54 P	2,4-Dinitrophenol	40.000	40.346	-0.9	106	0.00
55	Dibenzofuran	40.000	37.938	5.2	96	0.00
56 P	4-Nitrophenol	40.000	39.040	2.4	93	0.00
57	2,4-Dinitrotoluene	40.000	40.915	-2.3	98	0.00
58	Fluorene	40.000	37.361	6.6	95	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.521	3.7	96	0.00
60	Diethylphthalate	40.000	36.691	8.3	94	0.00
61	4-Chlorophenyl-phenylether	40.000	38.486	3.8	98	0.00
62	4-Nitroaniline	40.000	39.236	1.9	95	0.00
63	Azobenzene	40.000	36.276	9.3	91	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	94	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.514	-6.3	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.988	5.0	95	0.00
67	4-Bromophenyl-phenylether	40.000	38.881	2.8	98	0.00
68	Hexachlorobenzene	40.000	39.688	0.8	100	0.00
69	Atrazine	40.000	36.896	7.8	92	0.00
70 C	Pentachlorophenol	40.000	39.005	2.5	95	0.00
71	Phenanthrene	40.000	38.048	4.9	96	0.00
72	Anthracene	40.000	38.202	4.5	96	0.00
73	Carbazole	40.000	37.919	5.2	95	0.00
74	Di-n-butylphthalate	40.000	36.137	9.7	92	0.00
75 C	Fluoranthene	40.000	37.321	6.7	94	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	89	0.00
77	Benzydine	40.000	41.012	-2.5	98	0.00
78	Pyrene	40.000	39.298	1.8	94	0.00
79 S	Terphenyl-d14	80.000	77.284	3.4	95	0.00
80	Butylbenzylphthalate	40.000	39.621	0.9	92	0.00
81	Benzo(a)anthracene	40.000	38.389	4.0	90	0.00
82	3,3'-Dichlorobenzidine	40.000	37.227	6.9	87	0.00
83	Chrysene	40.000	38.355	4.1	91	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.515	6.2	86	0.00
85 c	Di-n-octyl phthalate	40.000	35.064	12.3	78	0.00
86 I	Perylene-d12	20.000	20.000	0.0	90	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.384	-3.5	102	0.00
88	Benzo(b)fluoranthene	40.000	38.390	4.0	87	0.00
89	Benzo(k)fluoranthene	40.000	38.536	3.7	89	0.00
90 C	Benzo(a)pyrene	40.000	38.157	4.6	87	0.00
91	Dibenzo(a,h)anthracene	40.000	41.799	-4.5	102	0.00
92	Benzo(g,h,i)perylene	40.000	42.202	-5.5	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042624\
Data File : BF137730.D
Acq On : 26 Apr 2024 11:47
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 26 12:28:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 24 03:57:12 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0