

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF050119\
 Data File : BF114072.D
 Acq On : 1 May 2019 16:32
 Operator : JU/SJ
 Sample : K2204-05 5X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TR-04-043019

Quant Time: May 02 05:37:21 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF042219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 23 03:32:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	84594	20.00	ng	0.00
21) Naphthalene-d8	8.18	136	289801	20.00	ng	0.00
39) Acenaphthene-d10	9.93	164	180686	20.00	ng	0.00
64) Phenanthrene-d10	11.42	188	335932	20.00	ng	0.00
76) Chrysene-d12	14.06	240	267454	20.00	ng	0.00
87) Perylene-d12	15.56	264	205405	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.52	112	83213	16.82	ng	0.00
7) Phenol-d6	6.52	99	128321	20.68	ng	0.00
23) Nitrobenzene-d5	7.45	82	60520	12.89	ng	0.00
42) 2,4,6-Tribromophenol	10.72	330	44232	20.10	ng	0.00
45) 2-Fluorobiphenyl	9.25	172	192214	16.64	ng	0.00
79) Terphenyl-d14	13.00	244	215210	16.50	ng	0.00
Target Compounds						
71) Phenanthrene	11.45	178	40784	2.416	ng	97
75) Fluoranthene	12.64	202	54253	2.945	ng	97
78) Pyrene	12.87	202	65118	3.482	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF050119\
Data File : BF114072.D
Acq On : 1 May 2019 16:32
Operator : JU/SJ
Sample : K2204-05 5X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TR-04-043019

Quant Time: May 02 05:37:21 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF042219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 23 03:32:35 2019
Response via : Initial Calibration

