

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF050120\
 Data File : BF120271.D
 Acq On : 1 May 2020 21:13
 Operator : MA/SJ
 Sample : L2463-07
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 NORTH-AOC14-CAR-WASH

Manual Integrations
 APPROVED

mohammad
 5/4/2020 3:45:34 PM

Quant Time: May 02 02:11:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF040820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 01 15:26:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.65	152	102900	20.00	ng	0.00
21) Naphthalene-d8	7.93	136	376695	20.00	ng	0.00
39) Acenaphthene-d10	9.69	164	163099	20.00	ng	-0.01
64) Phenanthrene-d10	11.17	188	280326	20.00	ng	0.00
76) Chrysene-d12	13.82	240	223489	20.00	ng	0.00
87) Perylene-d12	15.22	264	256432	20.00	ng	0.00

System Monitoring Compounds						
5) 2-Fluorophenol	5.26	112	645961	108.49	ng	0.00
7) Phenol-d6	6.30	99	823839	107.38	ng	-0.01
23) Nitrobenzene-d5	7.22	82	479697	65.88	ng	-0.01
42) 2,4,6-Tribromophenol	10.48	330	158364	79.31	ng	-0.01
45) 2-Fluorobiphenyl	9.02	172	862302	75.06	ng	0.00
79) Terphenyl-d14	12.76	244	798365	60.11	ng	-0.01

Target Compounds				Qvalue		
31) Naphthalene	7.96	128	135830	6.854	ng	96
37) 2-Methylnaphthalene	8.65	142	34660	2.580	ng	94
38) 1-Methylnaphthalene	8.75	142	26913	2.125	ng	# 95
49) Acenaphthylene	9.55	152	83237	5.161	ng	95
50) Dimethylphthalate	9.41	163	46492	3.685	ng	# 98
52) Acenaphthene	9.72	154	70911	6.591	ng	96
55) Dibenzofuran	9.89	168	64989	4.402	ng	98
58) Fluorene	10.23	166	94202	7.978	ng	95
71) Phenanthrene	11.20	178	787596	52.791	ng	98
72) Anthracene	11.25	178	266376	17.478	ng	97
73) Carbazole	11.41	167	54004	3.747	ng	97
75) Fluoranthene	12.39	202	999040	62.061	ng	97
78) Pyrene	12.62	202	1046654	55.553	ng	97
81) Benzo(a)anthracene	13.81	228	542691m	36.911	ng	
83) Chrysene	13.84	228	501422	33.564	ng	97
86) Indeno(1,2,3-cd)pyrene	16.57	276	270625	20.551	ng	95
88) Benzo(b)fluoranthene	14.83	252	639053m	34.642	ng	
89) Benzo(k)fluoranthene	14.85	252	215500m	13.539	ng	
90) Benzo(a)pyrene	15.17	252	479589	30.921	ng	99
91) Dibenzo(a,h)anthracene	16.58	278	67490m	4.912	ng	
92) Benzo(g,h,i)perylene	16.98	276	256900	18.943	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF050120\
Data File : BF120271.D
Acq On : 1 May 2020 21:13
Operator : MA/SJ
Sample : L2463-07
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
NORTH-AOC14-CAR-WASH

Manual Integrations
APPROVED

mohammad
5/4/2020 3:45:34 PM

Quant Time: May 02 02:11:39 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF040820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 01 15:26:36 2020
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.65 min Scan# 776
 Delta R.T. 0.00 min
 Lab File: BF120271.D
 Acq: 1 May 2020 21:13

Instrument :
 BNA_F
 ClientSampleId :
 NORTH-AOC14-CAR-WASH

Tot Ion	Ratio	Lower	Upper
152	100		
150	154.6	124.5	186.7
115	57.1	49.6	74.4




Manual Integrations
 APPROVED


mohammad
 5/4/2020 3:45:34 PM



#5
 2-Fluorophenol
 Concen: 108.488 ng
 RT: 5.26 min Scan# 540
 Delta R.T. 0.00 min
 Lab File: BF120271.D
 Acq: 1 May 2020 21:13

Tgt Ion	Ratio	Lower	Upper
112	100		
64	66.6	56.7	85.1
63	34.4	29.5	44.3





#7

Phenol-d6

Concen: 107.378 ng

RT: 6.30 min Scan# 717

Delta R.T. -0.01 min

Lab File: BF120271.D

Acq: 1 May 2020 21:13

Instrument :

BNA_F





ClientSampleId :

NORTH-AOC14-CAR-WASH

Tot Ion:	99	Resp:	823839
Ion Ratio	Lower	Upper	
99	100		
42	18.8	16.6	24.8
71	31.7	26.4	39.6

Manual Integrations
APPROVED

mohammad
5/4/2020 3:45:34 PM



#21

Naphthalene-d8

Concen: 20.000 ng




RT: 7.93 min Scan# 994


Delta R.T. -0.01 min

Lab File: BF120271.D

Acq: 1 May 2020 21:13

Tgt Ion:	136	Resp:	376695
Ion Ratio	Lower	Upper	
136	100		
137	10.5	8.7	13.1
54	11.0	9.1	13.7
68	6.0	4.6	6.8






#23
 Nitrobenzene-d5
 Concen: 65.879 ng
 RT: 7.22 min Scan# 872
 Delta R.T. -0.01 min
 Lab File: BF120271.D
 Acq: 1 May 2020 21:13

Instrument :
 BNA_F
 ClientSampleId :
 NORTH-AOC14-CAR-WASH




Tot Ion	Ratio	Lower	Upper
82	100		
128	45.4	36.3	54.5
54	60.7	50.6	75.8

Manual Integrations
 APPROVED

mohammad
 5/4/2020 3:45:34 PM




Abundance Ion 82.00 (81.70 to 82.70): BF1
 Ion 128.00 (127.70 to 128.70): E
 Ion 54.00 (53.70 to 54.70): BF1






#31
 Naphthalene
 Concen: 6.854 ng
 RT: 7.96 min Scan# 998
 Delta R.T. -0.01 min
 Lab File: BF120271.D
 Acq: 1 May 2020 21:13

Tgt Ion	Ratio	Lower	Upper
128	100		
129	10.1	9.5	14.3
127	12.6	11.3	16.9



Abundance Ion 128.00 (127.70 to 128.70): E
 Ion 129.00 (128.70 to 129.70): E
 Ion 127.00 (126.70 to 127.70): E









#37
2-Methylnaphthalene
Concen: 2.580 ng
RT: 8.65 min Scan# 1115
Delta R.T. -0.01 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Instrument :
BNA_F
ClientSampleId :
NORTH-AOC14-CAR-WASH

Tot Ion	Ratio	Lower	Upper
142	100		
141	82.4	71.7	107.5
115	33.6	27.9	41.9




Manual Integrations
APPROVED


mohammad
5/4/2020 3:45:34 PM



#38
1-Methylnaphthalene
Concen: 2.125 ng
RT: 8.75 min Scan# 1132
Delta R.T. -0.01 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Tgt Ion	Ratio	Lower	Upper
142	100		
141	86.9	73.7	110.5
116	6.0	3.0	4.6#









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.69 min Scan# 1292
Delta R.T. -0.01 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Instrument :
BNA_F
ClientSampleId :
NORTH-AOC14-CAR-WASH

Tot Ion	Ratio	Lower	Upper
164	100		
162	100.6	80.6	121.0
160	44.1	35.7	53.5




Manual Integrations
APPROVED


mohammad
5/4/2020 3:45:34 PM



#42
2,4,6-Tribromophenol
Concen: 79.305 ng
RT: 10.48 min Scan# 1427
Delta R.T. -0.01 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Tgt Ion	Ratio	Lower	Upper
330	100		
332	95.7	75.7	113.5
141	35.7	25.4	38.0






#45
2-Fluorobiphenyl
Concen: 75.064 ng
RT: 9.02 min Scan# 1178
Delta R.T. -0.01 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Instrument :
BNA_F
ClientSampleId :
NORTH-AOC14-CAR-WASH




Tot Ion	Ratio	Resp	Lower	Upper
172	100	862302		
171	35.6		29.6	44.4
170	23.4		20.2	30.2

Manual Integrations
APPROVED

mohammad
5/4/2020 3:45:34 PM




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#49
Acenaphthylene
Concen: 5.161 ng
RT: 9.55 min Scan# 1269
Delta R.T. -0.01 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Tgt Ion	Ratio	Resp	Lower	Upper
152	100	83237		
151	19.4		17.2	25.8
153	11.8		11.2	16.8



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E






#50
Dimethylphthalate
Concen: 3.685 ng
RT: 9.41 min Scan# 1245
Delta R.T. -0.02 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Instrument :
BNA_F
ClientSampleId :
NORTH-AOC14-CAR-WASH

Tot Ion	Ratio	Lower	Upper
163	100		
194	4.9	3.1	4.7#
164	9.8	8.2	12.2

Manual Integrations
APPROVED

mohammad
5/4/2020 3:45:34 PM






Abundance

Ion 163.00 (162.70 to 163.70): E


Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#52
Acenaphthene
Concen: 6.591 ng
RT: 9.72 min Scan# 1298
Delta R.T. -0.01 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Tgt Ion	Ratio	Lower	Upper
154	100		
153	114.0	88.1	132.1
152	51.0	43.0	64.4






Abundance

Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E









#55
Dibenzenofuran
Concen: 4.402 ng
RT: 9.89 min Scan# 1327
Delta R.T. -0.01 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Instrument :
BNA_F
ClientSampleId :
NORTH-AOC14-CAR-WASH

Tot Ion	Ratio	Lower	Upper
168	100		
139	35.7	30.1	45.1
169	13.6	11.0	16.6




Manual Integrations
APPROVED


mohammad
5/4/2020 3:45:34 PM



#58
Fluorene
Concen: 7.978 ng
RT: 10.23 min Scan# 1385
Delta R.T. -0.01 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Tgt Ion	Ratio	Lower	Upper
166	100		
165	99.4	75.4	113.2
167	13.5	11.2	16.8





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.17 min Scan# 1545

Delta R.T. -0.01 min

Lab File: BF120271.D

Acq: 1 May 2020 21:13

Instrument :

BNA_F


ClientSampleId :

NORTH-AOC14-CAR-WASH

Tot Ion:188	Resp:	280326
Ion Ratio	Lower	Upper
188 100		
94 8.1	7.5	11.3
80 8.1	9.0	13.4#

Manual Integrations
APPROVED




mohammad
5/4/2020 3:45:34 PM



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1



#71

Phenanthrene

Concen: 52.791 ng


RT: 11.20 min Scan# 1549

Delta R.T. -0.01 min

Lab File: BF120271.D

Acq: 1 May 2020 21:13



Tgt Ion:178	Resp:	787596
Ion Ratio	Lower	Upper
178 100		
176 19.2	16.2	24.4
179 15.8	13.2	19.8




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#72

Anthracene

Concen: 17.478 ng

RT: 11.25 min Scan# 1557

Delta R.T. -0.01 min

Lab File: BF120271.D

Acq: 1 May 2020 21:13

Instrument :

BNA_F





ClientSampleId :

NORTH-AOC14-CAR-WASH

Tot Ion:	178	Resp:	266376
Ion Ratio	Lower	Upper	
178	100		
176	18.5	16.1	24.1
179	14.9	13.0	19.4

Manual Integrations
APPROVED

mohammad
5/4/2020 3:45:34 PM



#73

Carbazole

Concen: 3.747 ng




RT: 11.41 min Scan# 1585


Delta R.T. -0.01 min

Lab File: BF120271.D

Acq: 1 May 2020 21:13

Tgt Ion:	167	Resp:	54004
Ion Ratio	Lower	Upper	
167	100		
166	21.6	18.2	27.4
139	12.0	10.7	16.1






#75
 Fluoranthene
 Concen: 62.061 ng
 RT: 12.39 min Scan# 1752
 Delta R.T. 0.00 min
 Lab File: BF120271.D
 Acq: 1 May 2020 21:13

Instrument :
 BNA_F
 ClientSampleId :
 NORTH-AOC14-CAR-WASH




Tot Ion	Ion	Ratio	Resp	Lower	Upper
202	100		999040		
101	10.5	0.0	32.2		
203	17.6	0.0	38.8		

Manual Integrations
 APPROVED

mohammad
 5/4/2020 3:45:34 PM




Abundance Ion 202.00 (201.70 to 202.70): E
 Ion 101.00 (100.70 to 101.70): E
 Ion 203.00 (202.70 to 203.70): E






#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 13.82 min Scan# 1995
 Delta R.T. -0.01 min
 Lab File: BF120271.D
 Acq: 1 May 2020 21:13

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
240	100		223489		
120	9.3	10.6	15.8#		
236	25.2	20.2	30.2		



Abundance Ion 240.00 (239.70 to 240.70): E
 Ion 120.00 (119.70 to 120.70): E
 Ion 236.00 (235.70 to 236.70): E









#78
Pvrene
Concen: 55.553 ng
RT: 12.62 min Scan# 1791
Delta R.T. 0.00 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Instrument :
BNA_F
ClientSampleId :
NORTH-AOC14-CAR-WASH

Tot Ion	Ratio	Lower	Upper
202	100		
200	20.0	16.8	25.2
203	17.6	15.1	22.7




Manual Integrations
APPROVED


mohammad
5/4/2020 3:45:34 PM



#79
Terphenyl-d14
Concen: 60.109 ng
RT: 12.76 min Scan# 1815
Delta R.T. -0.01 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.0	5.8	8.8
122	12.7	10.3	15.5









#81
Benzo(a)anthracene
Concen: 36.911 ng m
RT: 13.81 min Scan# 1993
Delta R.T. 0.00 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Instrument :
BNA_F
ClientSampleId :
NORTH-AOC14-CAR-WASH

Tot Ion	Ratio	Lower	Upper
228	100		
226	33.7	21.8	32.8
229	20.6	16.2	24.4




Manual Integrations
APPROVED


mohammad
5/4/2020 3:45:34 PM



#83
Chrysene
Concen: 33.564 ng
RT: 13.84 min Scan# 1999
Delta R.T. -0.01 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.8	25.4	38.0
229	22.0	16.6	24.8






#86
Indeno(1,2,3-cd)pyrene
Concen: 20.551 ng
RT: 16.57 min Scan# 2463
Delta R.T. -0.01 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Instrument :
BNA_F
ClientSampleId :
NORTH-AOC14-CAR-WASH




Tot Ion	Ratio	Lower	Upper
276	100		
138	21.1	19.6	29.4
277	23.8	20.0	30.0

Manual Integrations
APPROVED

mohammad
5/4/2020 3:45:34 PM




Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E






#87
Perylene-d12
Concen: 20.000 ng
RT: 15.22 min Scan# 2233
Delta R.T. 0.00 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Tgt Ion	Ratio	Lower	Upper
264	100		
260	23.3	18.7	28.1
265	22.9	17.7	26.5



Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E









#88
Benzo(b)fluoranthene
Concen: 34.642 ng m
RT: 14.83 min Scan# 2167
Delta R.T. 0.00 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Instrument :
BNA_F
ClientSampleId :
NORTH-AOC14-CAR-WASH

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.4	26.0
125	11.0	7.9	11.9




Manual Integrations
APPROVED


mohammad
5/4/2020 3:45:34 PM



#89
Benzo(k)fluoranthene
Concen: 13.539 ng m
RT: 14.85 min Scan# 2170
Delta R.T. -0.01 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.1	17.8	26.6
125	11.6	8.4	12.6









#90
Benzo(a)pyrene
Concen: 30.921 ng
RT: 15.17 min Scan# 2224
Delta R.T. 0.00 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Instrument :
BNA_F
ClientSampleId :
NORTH-AOC14-CAR-WASH

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.3	18.2	27.2
125	11.8	9.4	14.2




Manual Integrations
APPROVED


mohammad
5/4/2020 3:45:34 PM



#91
Dibenzo(a,h)anthracene
Concen: 4.912 ng m
RT: 16.58 min Scan# 2464
Delta R.T. -0.01 min
Lab File: BF120271.D
Acq: 1 May 2020 21:13

Tgt Ion	Ratio	Lower	Upper
278	100		
139	20.4	12.6	19.0#
279	30.3	19.0	28.6#






#92
 Benzo(a,h,i)perylene
 Concen: 18.943 ng
 RT: 16.98 min Scan# 2532
 Delta R.T. -0.01 min
 Lab File: BF120271.D
 Acq: 1 May 2020 21:13

Instrument :
 BNA_F
 ClientSampleId :
 NORTH-AOC14-CAR-WASH

Tot Ion	Ratio	Lower	Upper
276	100		
277	24.7	18.4	27.6
138	20.6	15.4	23.0

Manual Integrations
 APPROVED

mohammad
 5/4/2020 3:45:34 PM



Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

