

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050123\
 Data File : BF133154.D
 Acq On : 01 May 2023 14:20
 Operator : CG\JU
 Sample : 02507-01
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 HD-01-042623

Quant Time: May 01 15:03:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 01 15:02:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.728	152	190926	20.000	ng	0.00
21) Naphthalene-d8	8.016	136	757075	20.000	ng	0.00
39) Acenaphthene-d10	9.763	164	354025	20.000	ng	0.00
64) Phenanthrene-d10	11.245	188	545715	20.000	ng	0.00
76) Chrysene-d12	13.880	240	346483	20.000	ng	-0.01
86) Perylene-d12	15.304	264	339244	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.340	112	828000	65.512	ng	0.00
7) Phenol-d6	6.369	99	996825	63.542	ng	0.00
23) Nitrobenzene-d5	7.293	82	586197	41.794	ng	-0.01
42) 2,4,6-Tribromophenol	10.551	330	183785	51.619	ng	-0.01
45) 2-Fluorobiphenyl	9.092	172	1023364	48.361	ng	0.00
79) Terphenyl-d14	12.833	244	803662	40.003	ng	0.00
Target Compounds						Qvalue
9) Benzaldehyde	6.257	77	207	Below Cal	#	1
19) n-Nitroso-di-n-propyla...	7.293	70	77324	7.917	ng	# 77
25) Isophorone	7.293	82	584839	21.961	ng	# 63
51) 2,6-Dinitrotoluene	9.763	165	46109	8.299	ng	# 24
57) 2,4-Dinitrotoluene	9.763	165	46109	6.507	ng	# 23
63) Azobenzene	10.475	77	55571	2.250	ng	# 1
67) 4-Bromophenyl-phenylether	10.551	248	12198	2.061	ng	# 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050123\
Data File : BF133154.D
Acq On : 01 May 2023 14:20
Operator : CG\JU
Sample : 02507-01
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
HD-01-042623

Quant Time: May 01 15:03:20 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 01 15:02:31 2023
Response via : Initial Calibration

