

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050324\
 Data File : BF137831.D
 Acq On : 03 May 2024 16:19
 Operator : CG\JU
 Sample : P2354-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 V359

Quant Time: May 03 16:45:07 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 01 03:25:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.051	152	145199	20.000	ng	0.00
21) Naphthalene-d8	8.333	136	573104	20.000	ng	-0.01
39) Acenaphthene-d10	10.098	164	331212	20.000	ng	-0.01
64) Phenanthrene-d10	11.592	188	510812	20.000	ng	# 0.00
76) Chrysene-d12	14.239	240	258139	20.000	ng	-0.01
86) Perylene-d12	15.809	264	259687	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.687	112	986326	113.733	ng	0.01
7) Phenol-d6	6.669	99	1102785	100.254	ng	0.00
23) Nitrobenzene-d5	7.616	82	844093	85.570	ng	0.00
42) 2,4,6-Tribromophenol	10.892	330	441671	130.604	ng	0.00
45) 2-Fluorobiphenyl	9.410	172	1746710	85.558	ng	-0.01
79) Terphenyl-d14	13.180	244	1579847	102.565	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050324\
Data File : BF137831.D
Acq On : 03 May 2024 16:19
Operator : CG\JU
Sample : P2354-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
V359

Quant Time: May 03 16:45:07 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 03:25:08 2024
Response via : Initial Calibration

