

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050422\
 Data File : BF128062.D
 Acq On : 04 May 2022 10:28
 Operator : CG\JU
 Sample : PB144541BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB144541BL

Quant Time: May 05 06:00:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 04:23:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.916	152	108188	20.000	ng	0.00
21) Naphthalene-d8	8.198	136	433382	20.000	ng	0.00
39) Acenaphthene-d10	9.957	164	251910	20.000	ng	0.00
64) Phenanthrene-d10	11.451	188	462595	20.000	ng	0.00
76) Chrysene-d12	14.098	240	329796	20.000	ng	0.00
86) Perylene-d12	15.604	264	260729	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.551	112	819070	120.586	ng	0.01
7) Phenol-d6	6.545	99	1067361	121.104	ng	0.00
23) Nitrobenzene-d5	7.481	82	786134	85.299	ng	0.00
42) 2,4,6-Tribromophenol	10.757	330	324607	128.005	ng	0.00
45) 2-Fluorobiphenyl	9.275	172	1318655	88.556	ng	0.00
79) Terphenyl-d14	13.045	244	1544456	85.633	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050422\
Data File : BF128062.D
Acq On : 04 May 2022 10:28
Operator : CG\JU
Sample : PB144541BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB144541BL

Quant Time: May 05 06:00:31 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 04 04:23:29 2022
Response via : Initial Calibration

