

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050422\
 Data File : BF128061.D
 Acq On : 04 May 2022 09:59
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 05 05:59:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 04:23:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	73	0.00
2	1,4-Dioxane	0.599	0.585	2.3	70	0.02
3	Pyridine	1.535	1.574	-2.5	71	0.02
4	n-Nitrosodimethylamine	0.743	0.796	-7.1	76	0.02
5 S	2-Fluorophenol	1.256	1.254	0.2	73	0.00
6	Aniline	1.936	2.014	-4.0	73	0.00
7 S	Phenol-d6	1.629	1.686	-3.5	75	0.00
8	2-Chlorophenol	1.303	1.333	-2.3	73	0.00
9	Benzaldehyde	1.008	0.904	10.3	74	0.00
10 C	Phenol	1.643	1.708	-4.0	75	0.00
11	bis(2-Chloroethyl)ether	1.363	1.391	-2.1	74	0.00
12	1,3-Dichlorobenzene	1.500	1.501	-0.1	73	0.00
13 C	1,4-Dichlorobenzene	1.497	1.501	-0.3	73	0.00
14	1,2-Dichlorobenzene	1.386	1.394	-0.6	74	0.00
15	Benzyl Alcohol	1.108	1.412	-27.4#	86	0.00
16	2,2'-oxybis(1-Chloropropane	2.083	2.117	-1.6	72	0.00
17	2-Methylphenol	1.124	1.134	-0.9	70	0.00
18	Hexachloroethane	0.550	0.580	-5.5	76	0.00
19 P	n-Nitroso-di-n-propylamine	0.997	1.063	-6.6	77	0.00
20	3+4-Methylphenols	1.358	1.404	-3.4	73	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	72	0.00
22	Acetophenone	0.487	0.501	-2.9	75	0.00
23 S	Nitrobenzene-d5	0.425	0.469	-10.4	78	0.00
24	Nitrobenzene	0.410	0.443	-8.0	76	0.00
25	Isophorone	0.716	0.761	-6.3	75	0.00
26 C	2-Nitrophenol	0.175	0.198	-13.1	77	0.00
27	2,4-Dimethylphenol	0.290	0.296	-2.1	70	0.00
28	bis(2-Chloroethoxy)methane	0.459	0.481	-4.8	75	0.00
29 C	2,4-Dichlorophenol	0.304	0.315	-3.6	73	0.00
30	1,2,4-Trichlorobenzene	0.345	0.359	-4.1	75	0.00
31	Naphthalene	1.019	1.033	-1.4	73	0.00
32	Benzoic acid	0.208	0.205	1.4	64	0.00
33	4-Chloroaniline	0.403	0.415	-3.0	72	0.00
34 C	Hexachlorobutadiene	0.217	0.242	-11.5	80	0.00
35	Caprolactam	0.098	0.097	1.0	67	0.00
36 C	4-Chloro-3-methylphenol	0.325	0.348	-7.1	75	0.00
37	2-Methylnaphthalene	0.673	0.692	-2.8	74	0.00
38	1-Methylnaphthalene	0.654	0.669	-2.3	74	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	72	0.00
40	1,2,4,5-Tetrachlorobenzene	0.590	0.630	-6.8	77	0.00
41 P	Hexachlorocyclopentadiene	0.319	0.327	-2.5	69	0.00
42 S	2,4,6-Tribromophenol	0.201	0.227	-12.9	81	0.00
43 C	2,4,6-Trichlorophenol	0.388	0.414	-6.7	75	0.00
44	2,4,5-Trichlorophenol	0.422	0.440	-4.3	72	0.00
45 S	2-Fluorobiphenyl	1.182	1.329	-12.4	79	0.00
46	1,1'-Biphenyl	1.481	1.553	-4.9	75	0.00
47	2-Chloronaphthalene	1.128	1.150	-2.0	73	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050422\
 Data File : BF128061.D
 Acq On : 04 May 2022 09:59
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 05 05:59:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 04:23:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.381	0.424	-11.3	76	0.00
49	Acenaphthylene	1.725	1.764	-2.3	73	0.00
50	Dimethylphthalate	1.341	1.383	-3.1	73	0.00
51	2,6-Dinitrotoluene	0.290	0.302	-4.1	71	0.00
52 C	Acenaphthene	1.160	1.194	-2.9	73	0.00
53	3-Nitroaniline	0.322	0.314	2.5	66	0.00
54 P	2,4-Dinitrophenol	0.153	0.169	-10.5	72	0.00
55	Dibenzofuran	1.673	1.688	-0.9	72	0.00
56 P	4-Nitrophenol	0.246	0.220	10.6	59	0.00
57	2,4-Dinitrotoluene	0.385	0.392	-1.8	69	0.00
58	Fluorene	1.249	1.305	-4.5	76	0.00
59	2,3,4,6-Tetrachlorophenol	0.355	0.362	-2.0	71	0.00
60	Diethylphthalate	1.315	1.359	-3.3	73	0.00
61	4-Chlorophenyl-phenylether	0.620	0.656	-5.8	77	0.00
62	4-Nitroaniline	0.315	0.301	4.4	65	0.00
63	Azobenzene	1.447	1.494	-3.2	73	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	68	0.00
65	4,6-Dinitro-2-methylphenol	0.119	0.143	-20.2	72	0.00
66 c	n-Nitrosodiphenylamine	0.622	0.653	-5.0	72	0.00
67	4-Bromophenyl-phenylether	0.222	0.241	-8.6	74	0.00
68	Hexachlorobenzene	0.234	0.257	-9.8	75	0.00
69	Atrazine	0.192	0.199	-3.6	71	0.00
70 C	Pentachlorophenol	0.137	0.132	3.6	61	0.00
71	Phenanthrene	1.047	1.068	-2.0	70	0.00
72	Anthracene	1.043	1.077	-3.3	71	0.00
73	Carbazole	1.005	0.995	1.0	68	0.00
74	Di-n-butylphthalate	1.130	1.190	-5.3	71	0.00
75 C	Fluoranthene	1.102	1.091	1.0	68	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	66	0.00
77	Benzidine	0.378	0.478	-26.5#	77	0.00
78	Pyrene	1.614	1.668	-3.3	68	0.00
79 S	Terphenyl-d14	1.094	1.195	-9.2	74	0.00
80	Butylbenzylphthalate	0.633	0.708	-11.8	71	0.00
81	Benzo(a)anthracene	1.333	1.391	-4.4	68	0.00
82	3,3'-Dichlorobenzidine	0.423	0.414	2.1	64	0.00
83	Chrysene	1.273	1.277	-0.3	66	0.00
84	Bis(2-ethylhexyl)phthalate	0.839	0.949	-13.1	72	0.00
85 c	Di-n-octyl phthalate	1.300	1.465	-12.7	71	0.00
86 I	Perylene-d12	1.000	1.000	0.0	65	0.00
87	Indeno(1,2,3-cd)pyrene	1.335	1.397	-4.6	64	0.00
88	Benzo(b)fluoranthene	1.256	1.386	-10.4	71	0.00
89	Benzo(k)fluoranthene	1.266	1.230	2.8	60	0.00
90 C	Benzo(a)pyrene	1.042	1.090	-4.6	65	0.00
91	Dibenzo(a,h)anthracene	1.071	1.142	-6.6	65	0.00
92	Benzo(g,h,i)perylene	1.128	1.158	-2.7	62	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050422\
Data File : BF128061.D
Acq On : 04 May 2022 09:59
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 05 05:59:52 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 04 04:23:29 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0