



Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF050422\  
 Data File : BF128060.D  
 Acq On : 04 May 2022 09:30  
 Operator : CG\JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF042222.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed May 04 04:23:29 2022






AutoFind: Scans 1598, 1599, 1600; Background Corrected with Scan 1592

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 38.4         | 61189      | PASS                |
| 68             | 69              | 0.00            | 2               | 1.6          | 987        | PASS                |
| 69             | 198             | 0.00            | 100             | 37.7         | 60040      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.7          | 415        | PASS                |
| 127            | 198             | 10              | 80              | 46.1         | 73459      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.5          | 792        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 159451     | PASS                |
| 199            | 198             | 5               | 9               | 6.5          | 10349      | PASS                |
| 275            | 198             | 10              | 60              | 28.1         | 44813      | PASS                |
| 365            | 198             | 1               | 100             | 4.1          | 6465       | PASS                |
| 441            | 198             | 0.01            | 100             | 16.1         | 25645      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 163416     | PASS                |
| 443            | 442             | 15              | 24              | 19.4         | 31744      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF050422\  
Data File : BF128060.D  
Acq On : 04 May 2022 09:30  
Operator : CG\JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Quant Time: May 04 13:48:50 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF042222.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed May 04 04:23:29 2022  
Response via : Initial Calibration



TIC: BF128060.D\data.ms

**(70) Pentachlorophenol (C)**

11.251min (-0.006) 22274.88 ng




response 127341

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 66.30  | 64.48  |
| 264.00 | 66.40  | 62.26  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF050422\  
 Data File : BF128060.D  
 Acq On : 04 May 2022 09:30  
 Operator : CG\JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Quant Time: May 04 13:48:50 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF042222.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed May 04 04:23:29 2022  
 Response via : Initial Calibration



TIC: BF128060.D\data.ms

**(77) Benzidine**

12.792min ( 0.000) 546046.42 ng m

response 671579

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.10  | 14.21  |
| 183.00 | 11.20  | 11.71  |
| 0.00   | 0.00   | 0.00   |

# DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 5/4/2022      | BNA_F            | <u>BF128060.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 386000           | 13.61              |
| DDD           | 7459             | 13.31              |
| DDE           | 919              | 12.974             |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 8378          | 394378           | 2.12               |
|               |                  |                    |

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP